

Declaration of Conformity



Type of equipment: breathalyzer
Brand Name /Trade Mark: Alcofind
Type designation /model: AF-35, AF-30, AF-33, ELITE and PRO X-5
Manufacturer: DA Tech Co., Ltd.

In accordance with the following Directives:

2014/30/EU The Electromagnetic Compatibility Directive

The following harmonized European standards and technical specifications have been applied:

EN 61326-1:2013	Electrical equipment for measurement, control and laboratory use - EMC requirements - Part 1: General requirements
EN 55011:2009 +A1:2010	Industrial, scientific and medical (ISM) radio-frequency equipment - Electromagnetic disturbance characteristics - Limits and methods of measurement
EN 61326-2-2:2006	Electrical equipment for measurement, control and laboratory use - EMC requirements -- Part 2-2: Particular requirements - Test configurations, operational conditions and performance criteria for portable test, measuring and monitoring equipment used in low-voltage distribution systems
EN 61000-4-2:2009	Electrostatic discharge immunity test
EN 61000-4-3:2006 +A1:2008 +A2:2010	Radiated, radio-frequency, electromagnetic field immunity test
EN 61000-4-8:2010	Electromagnetic compatibility (EMC) - Part 4-8: Testing and measurement techniques - Power frequency magnetic field immunity test

Test report issued by:

EMC: CTK Co., Ltd.

The CE Marking on the products and/or their packaging signifies that DA Tech Co., Ltd. hold the reference technical file available to the European Union authorities.

Place and date of issue:

39, Pyeongcheon-ro 141 beon-fil, Bupyeong-gu, Incheon, Korea

/ 2014-07-17

Authorized Signatory:

Name: KWANG HEE, PARK

Title: CEO

Signature: