



KERN & Sohn GmbH

Ziegelei 1

D-72336 Balingen

email: info@kern-sohn.com

Phone: +49-[0]7433- 9933-0

Fax: +49-[0]7433-9933-149

Internet: www.kern-sohn.com

Operating instructions Density Determination Set

KERN PBS-A03/A04

Version 1.4

07/2016

GB



PBS-A03/A04-BA-e-1614



KERN PBS-A03/A04

Version 1.4 07/2016

Operating instructions

Set for density determination for precision balances

KERN PBJ/PBS

Contents:

1	INTRODUCTION	3
1.1	SCOPE OF SUPPLY	3
1.2	DIMENSIONS.....	5
2	INSTALLING THE DENSITY DETERMINATION SET	6
3	PRINCIPLE OF DENSITY DETERMINATION.....	8
3.1	INFLUENCING MAGNITUDES AND ERROR SOURCES	9
4	DENSITY DETERMINATION OF SOLIDS.....	10
☞	ACCESS MODE FOR THE DENSITY DETERMINATION OF SOLIDS.	10
☞	INPUT „DENSITY AUXILIARY FLUID“	12
☞	ACCESS MODE FOR THE DENSITY DETERMINATION OF SOLIDS.	14
5	DETERMINING DENSITY OF LIQUIDS	17
5.1	VOLUME DETERMINATION OF THE SINKER.....	17
5.2	DENSITY DETERMINATION AT KNOWN VOLUME OF THE SINKER	19
☞	MODE TO CALL UP DENSITY DETERMINATION OF LIQUIDS	19
☞	ENTER SINKER VOLUME.....	21
☞	DENSITY DETERMINATION OF THE TEST FLUID	23
6	PRECONDITIONS FOR PRECISE MEASUREMENTS.....	26
6.1	CALCULATION OF RESULTS	26
6.2	INFLUENCE FACTORS FOR MEASUREMENT ERRORS	27
6.2.1	Air bubbles	27
6.2.2	Solid body sample	27
6.2.3	Liquids.....	27
6.2.4	Surface.....	27
6.3	GENERAL INFORMATION.....	27
6.3.1	Density / Relative Density	27
7	DENSITY TABLE FOR FLUIDS.....	28
8	USER INSTRUCTIONS.....	29

1 Introduction

KERN PBS-A03	KERN PBS-A04
• Set for density determination for precision balances of the series KERN PBJ/PBS with big weighing plate (180 x 170 mm).	• Set for density determination for precision balances of the series KERN PBJ/PBS with small weighing plate (105 x 105 mm).
• When using the density set, the capacity of the balance is reduced by approx. 100 g.	• When using the density set, the capacity of the balance is reduced by approx. 290 g.



- In order to guarantee a safe and trouble-free operation, please read carefully the operating instructions.
- These operating instructions only describe the operation of the density determination set. For further information on how to operate your balance please refer to the operating instructions supplied with each balance.

1.1 Scope of supply



Fig. 1: Installed density set **KERN PBS-A04**

1. Weighing tray holder



2. Combination weighing tray



3. Container table



4. Container




5. Weighing tray carrier,
4 items



6. Glass sinker



1.2 Dimensions



2 Installing the density determination set

i

- If necessary, carry out necessary adjustment before installation of the density set.
- When the density set is installed, correct adjustment is not possible.
- For reasons of adjustment, take away the density set and place the standard weighing plate.
- The following pictures show the density set **KERN PBS-A03** at a balance with big weighing plate. The density set **KERN PBS-A04** must be installed at the same way.


1. Switch off balance and separate it from the power supply.

2. Remove standard weighing plate.



3. Remove carrier of the standard weighing plate and replace by weighing tray carrier of the density set.



4. Put balance tray holder according to fig. into the four balance tray carriers. Thereby observe the right position, the openings on the upper side must point forwards.



5. Place the container table in a way that it does not touch the balance tray holder.



6. Put the containers in the center of the container table



7. Hook up the combination balance tray according to illustration; ensure that it does not touch the container.



3 Principle of Density Determination

Three physical magnitudes are the **volume** and the **mass** of bodies as well as the **density** of matter. In density mass and volume are related.

Density [ρ] is the relation of mass [m] to volume [V].

$$\rho = \frac{m}{V}$$

SI-unit of density is kilogram divided by cubic meter (kg/m^3). 1 kg/m^3 equals the density of a homogenous body that, for a mass of 1 kg , has the volume of 1 m^3 .

Additional frequently applied units include:

$$1 \frac{\text{g}}{\text{cm}^3}, 1 \frac{\text{kg}}{\text{m}^3}, 1 \frac{\text{g}}{\text{l}}$$

The application of this density determination set in combination with the KERN PBS/PBJ balances provides fast and safe determination of solids and fluids. Our set uses the "**Principle of Archimedes**" to determine density:

BUOYANCY IS A FORCE. IT AFFECTS A BODY THAT IS IMMERSSED INTO A FLUID. THE BUOYANCY OF THE BODY EQUALS THE WEIGHT FORCE OF THE DISPLACED FLUID. THE FORCE OF BUOYANCY ACTS VERTICALLY UPWARDS.

Thus, density is calculated according to the formulae below:

Determining density of solid bodies

Our balances enable weighing of solids in air [A] as well as water [B]. If the density of the buoyancy medium is known [ρ_0] the density of the solid [ρ] is calculated as follows:

$$\rho = \frac{A}{A-B} \rho_0$$

ρ = density of sample

A = weight of the sample in air

B = weight of sample in measuring fluid

ρ_0 = density of measuring fluid

Determining density of liquids

The density of a fluid is determined with the help of a sinker providing a known volume [V]. The sinker is weighed in air [A] as well as in the test fluid [B].

According to the Archimedes' Principle a body immersed in a fluid experiences a force of buoyancy. [G]. This force equals the weight force of the fluid displaced by the volume of the body.

The volume [V] of the immersed body equals the volume of the displaced fluid.

$$\rho = \frac{G}{V}$$

G = buoyancy of sinker

Buoyancy of sinker =

Weight of the sinker in air [A] - weight sinker in test liquid [B]

From this follows:

$$\rho = \frac{A-B}{V} + \rho_L$$

ρ = Density of test liquid

A = weight of sinker in air

B = weight of sinker in sample fluid

V = volume of sinker*

ρ_L = Air density (0.0012 g/cm³)

* If the volume of the sinker is unknown, this can be determined e.g. in water and be calculated as follows, see chap. 5.1.

$$V = \frac{A-B}{\rho_w}$$

V = volume of sinker

A = weight of sinker in air

B = weight of sinker in water

ρ_w = density of water

3.1 Influencing magnitudes and error sources

- ⇒ Air pressure
- ⇒ temperature
- ⇒ Volume deviance of the sinker ($\pm 0,005 \text{ cm}^3$)
- ⇒ Surface tension of the liquid
- ⇒ Air bubbles
- ⇒ Immersion depth of the sample dish of sinker
- ⇒ Porosity of the solid

4 Density determination of solids

For the determination of the density of solids, the solid is first weighed in air and then in the aid liquid, whose density is known. From the weight difference results the buoyancy from where the software calculates the density.

As aid liquid, usually distilled water or ethanol is used, see density tables chapter 7.



- ⇒ Prepare balance as described in chapter 2 "Installation of density determination set".
- ⇒ Fill auxiliary fluid into the container. Filling height should be approx. $\frac{3}{4}$ of the capacity. Bring liquid and instruments to the right temperature until you achieve a constant temperature. Observe the warm-up time of the balance.

Access mode for the density determination of solids.

- ⇒ Switch on balance



- ⇒ In weighing mode press  repeatedly until „U“ flashes.



⇒ Press .




⇒ Press  repeatedly until „U-▼d“ is displayed.
From now on the balance is in density determining mode for solids.



⇒ Press , the current setting of „SG HOLD“ will be displayed.

⇒ Press  to select the desired setting for the display mode.

Every time  is pressed the symbol (→) is unhidden or hidden.

(→) **on**: Hold display mode (the displayed density value will be kept on display).

(→) **off**: Continuous display mode (the displayed density value will be continuously updated according to the weight change in the auxiliary fluid).

 Input „Density auxiliary fluid“



- ⇒ Press , the density of the auxiliary fluid saved as last will be displayed, at the first input, the value will be zero. The active digit is flashing. When changing by the arrow keys, enter first the numeric value of the density considering the current temperature (see chap. 7), then set the decimal place.



Use  to increase the numeric value of the flashing cipher.

Use  to move the cipher selection to the right, the respective active position flashes.

Input example for the value „1.000“:

- ⇒ Press  repeatedly until the forth digit flashes.

Press  to set the numeric value „1“.




- ⇒ To set the decimal place , take the last place using , when this place dot is flashing, press  anew. The ▼ symbol pops up.



⇒ Use  to define the position of the decimal place.



⇒ Confirm input by . Observe that the stability mark is displayed, if not, the entry has not been accepted.



⇒ Press  repeatedly or keep pressed for 3 sec until the gram display will appear.



 Access mode for the density determination of solids.

1. In weighing mode press  repeatedly until [Air ▼ g d] will be displayed. After approx. 2 sec the display changes automatically into the density determination mode for solids.



Reset to zero if necessary by pressing .

2. Place sample in the upper sample dish.




3. The weight of the „sample in air“ is displayed. When an optional printer is connected, the displayed value can be edited using .


Wait for stability display (➡), then press .




4. For approx. 2 s “wAtr” will be displayed. The display changes automatically to [0.00 ▼ g d] Remove sample from the upper sample dish and set to zero if necessary.



5. Remove sample from the upper sample dish. Reset to zero if necessary by pressing



6. Lay the sample into the lower sample dish and immerse it in the auxiliary liquid. Make sure that the sample is at least 1 cm immersed and has, if possible, no air bubbles adhering to it.




7. The weight of the „sample in auxiliary liquid“ is displayed. When an optional printer is

connected, the displayed value can be edited using .



8. Wait for stability display (➡), then press . The scale determines the density of the sample and then displays the result. When „Hold display mode“ is enabled, the symbol [*] will be unhidden.




9. When connecting an optional printer, the result will be printed after pressing



Printout example:

3.4999DS

If necessary, remove the decimal place by pressing the UNIT-button as follows:



- ⇒ **To start the next measurement** press , before the current sample is removed from the lower sample dish.
For the new measurement start at step 2.

- ⇒ Using  the balance returns into the weighing mode.



To avoid damage by corrosion, don't leave the immersion basket set immersed in liquid for a longer time.

5 Determining density of liquids

For density determination of liquids, a plummet is used, whose volume is known. The sinker is weighed first in air and then in the liquid whose density is to be determined. From the weight difference results the buoyancy from where the software calculates the density.

The volume of the added glass sinker will be computed as described in the following chapter 2.1.

5.1 Volume determination of the sinker




- ⇒ Prepare balance as described in chapter 2 "Installation of density determination set".
- ⇒ Dest. Fill water into the container. Filling height should be approx. $\frac{3}{4}$ of the capacity. Heat sample liquid until temperature is constant.
- ⇒ Keep the sinker ready


1. Switch-on balance, repeatedly press the  key until the balance is in weighing mode.



2. Place sinker in the upper sample dish. Wait for stability display to appear; write down the displayed weighing value.



3. Place sinker in the lower sample dish. Wait for stability display to appear; write down the displayed weighing value.



The volume of the sinker is computed by applying the following formula.

$$V = \frac{A - B}{\rho_w}$$

V = volume of sinker

A = weight of the sinker in air = 99.998 g

B = weight of the sinker in water = 87.607 g

ρ_w = density of water (see chap. 7) at 20°C = 0.9982 g/cm³

$$V = \frac{99.998 \text{ g} - 87.607 \text{ g}}{0.9982 \text{ g/cm}^3} = 12.413 \text{ cm}^3$$

5.2 Density determination at known volume of the sinker



- ⇒ Prepare the balance as described in chapter 2 „Installing density determination set“ (step 1-6).
- ⇒ Keep the sinker ready
- ⇒ Fill test liquid in the container. Filling height should be approx. $\frac{3}{4}$ of the capacity.
- ⇒ Control the temperature of the liquid, instruments and the displacement body until the temperature is constant. Observe the warm-up time of the balance.

Mode to call up density determination of liquids

- ⇒ Switch on balance




- ⇒ In weighing mode press  repeatedly until „U“ flashes.



⇒ Press .



⇒ Press  repeatedly until [U-d] is displayed.
From now on the balance is in density determining mode for liquids.




⇒ Press , the current setting of „SG HOLD“ will be displayed.

⇒ Press  to select the desired setting for the display mode.

Every time  is pressed the symbol (→) is unhidden or hidden.

(→) **on**: Hold display mode (the displayed density value will be kept on display).

(→) **off**: Continuous display mode (the displayed density value will be continuously updated according to the weight change in the auxiliary fluid).



Enter sinker volume



- ⇒ Press , the volume of the sinker saved as last will be displayed, at the first input, the value will be zero. The active digit is flashing. When changing by the arrow keys, enter first the numeric value of the volume (see chapter 2.1), then set the decimal place.



Use  to increase the numeric value of the flashing cipher.

Use  to move the cipher selection to the right, the respective active position flashes.

Input example for the value „15.127“:

- ⇒ Press  repeatedly until the fifth digit flashes.

Press  to set the numeric value „1“.



Press  to select the following cipher and press  to set the numeric value „5“. Enter all ciphers in the same way.



- ⇒ To set the decimal place, take the last place using , when this place dot is flashing, press  anew. The ▼ symbol pops up.



⇒ Use  to define the position of the decimal place.



⇒ Acknowledge with . Observe that the stability mark is displayed, if not, the entry has not been accepted.



⇒ Press  repeatedly or keep pressed for 3 sec until the gram display will appear.



Density determination of the test fluid

1. In weighing mode press  repeatedly until [Air g d ►] will be displayed. After approx. 2 sec the display changes automatically into the density determination mode for liquids. Reset to zero if necessary by pressing .



2. Hang the sinker on the suspension device





3. The weight of the „sinker in air“ is displayed.
When an optional printer is connected, the displayed value can be edited using the PRINT button.

Wait for stability display (►), then press .



4. For approx. 2 s "wAtr" will be displayed.
The display changes automatically to "g d ▶"



5. Remove the glass sinker. Reset to zero if necessary by pressing .




6. Put glass container with test fluid onto the platform. Suspend again the glass sinker and immerse it completely into the liquid. The weight of the „sinker in liquid“ is displayed.




When an optional printer is connected, the displayed value can be edited using the PRINT button.

7. Wait for stability display (▀), then press . The scale determines the density of the liquid and then displays the result.
When „Hold display mode“ is enabled, the symbol [*] will be unhidden.




8. When connecting an optional printer, the result will be printed after pressing .

Printout example:



If necessary, remove the decimal place by pressing the UNIT-button as follows:



⇒ **To start the next measurement** press , before the sinker will be removed.
For the new measurement start at step 2.

⇒ Using  the balance returns into the weighing mode.

6 Preconditions for Precise Measurements

There are numerous error possibilities during density determination. Accurate knowledge and caution are required to achieve precise results when applying this density set in combination with the balance.

6.1 Calculation of Results

The balance displays results for density determination by giving four decimal places. However, this does not mean that the results are accurate down to the last decimal place as this would be the case for a calculated value. Therefore all weighing results used for calculations have to be examined closely.

Example for density determination of solids:

To ensure high-grade results, numerators as well as common denominators of the formula below must show the desired accuracy. If either of them is instable or flawed, the result, too, will be instable or flawed.

$$\rho = \frac{A}{A-B} \rho_0$$

ρ = density of sample
 A = weight of the sample in air
 B = weight of sample in measuring fluid
 ρ_0 = density of measuring fluid

The use of a heavy sample contributes to the accuracy of a result. this increases the numerical value. The use of a light-weight sample, too, contributes to the accuracy of a result because this increases buoyancy (A-B). As a consequence, the result of the common denominator increases. Bear also in mind that the accuracy of the density of the measuring fluid ρ_0 enters into the common denominator and, thus, has considerable influence on the accuracy of the result.

The result for the density of the sample cannot be more accurate than the least accurate of the aforementioned individual entities.

6.2 Influence Factors for Measurement Errors

6.2.1 Air bubbles

A small bubble of, for example, 1mm^3 will have a considerable influence on the measurement if the sample is small. Buoyancy will be increased by approximately 1mg resulting immediately in an error of 2 digits. Hence, it has to be ensured that no air bubbles cling to the solid immersed in the fluid. The same applies to the sinker that is immersed in the test fluid.

Take great care when removing air bubbles by swirling, to prevent the fluid from spurting out and splashing onto the sifting bowl or from water splashing. Moisture on the suspension bracket of the sifting bowl results in increased weight.

Do not touch the solid sample or sinker with bare fingers. An oily surface causes air bubbles when immersing the specimen in fluids.

Do not place solid samples (in particular flat objects) in the sifting bowls outside the liquid as this would result in air bubbles when immersed together. For this reason examine the bottom of the sifting bowl for air bubbles after the specimen had been immersed in fluid.

6.2.2 Solid body sample

A sample possessing too great a volume that is immersed in fluid will result in an increase in fluid level inside the glass pitcher. As a result, part of the suspension bracket of the sifting bowl will also be immersed causing buoyancy to increase. As a consequence the weight of the specimen in the fluid will drop.

Samples that change the volume or assimilate fluid are unsuitable for measurement.

6.2.3 Liquids

Water temperature is another factor to be taken into consideration. The density of water changes by c. 0.01% per degree Celsius. A temperature measurement showing an error of 1 degree Celsius results in an inaccurate fourth decimal place.

6.2.4 Surface

The suspension bracket of the sifting bowl penetrates the surface of the fluid. This state undergoes continuous change. If the sample or the sinker is relatively small, the surface tension will impair repeatability. The addition of a small amount of detergent makes the surface tension negligible and increases repeatability.

6.3 General information

6.3.1 Density / Relative Density

Relative density follows from the weight of a specimen divided by the weight of water (at 4° Celsius) of the same volume. For this reason relative density does not have a unit.

Density equals mass divided by volume.

The application of the relative density instead of the density of a fluid in a formula produces an incorrect result. In the case of fluids only their density is physically meaningful.

7 Density Table for Fluids

Temperature [°C]	Density ρ [g/cm ³]		
	Water	Ethyl alcohol	Methyl alcohol
10	0.9997	0.7978	0.8009
11	0.9996	0.7969	0.8000
12	0.9995	0.7961	0.7991
13	0.9994	0.7953	0.7982
14	0.9993	0.7944	0.7972
15	0.9991	0.7935	0.7963
16	0.9990	0.7927	0.7954
17	0.9988	0.7918	0.7945
18	0.9986	0.7909	0.7935
19	0.9984	0.7901	0.7926
20	0.9982	0.7893	0.7917
21	0.9980	0.7884	0.7907
22	0.9978	0.7876	0.7898
23	0.9976	0.7867	0.7880
24	0.9973	0.7859	0.7870
25	0.9971	0.7851	0.7870
26	0.9968	0.7842	0.7861
27	0.9965	0.7833	0.7852
28	0.9963	0.7824	0.7842
29	0.9960	0.7816	0.7833
30	0.9957	0.7808	0.7824
31	0.9954	0.7800	0.7814
32	0.9951	0.7791	0.7805
33	0.9947	0.7783	0.7896
34	0.9944	0.7774	0.7886
35	0.9941	0.7766	0.7877

8 User Instructions

- To form a reproducible mean value several density measurement are necessary
- Remove fat from solvent-resistant sample / sinker /beaker.
- Regularly clean sample dishes/ sinker/beaker, do not touch immersed part with your hands
- Dry sample/ sinker/pincers after each measurement.
- Adjust sample size to sample dish (ideal sample size > 5 g).
- Only use distilled water.
- When immersing for the first time, lightly shake sample dishes and sinker, in order to Dissolve air bubbles.
- Always ensure that, when re-immersing into the liquid no additional bubbles adhere; it is better to use pincers to place the sample.
- Remove firmly adherent air bubbles with a fine brush or a similar tool.
- To avoid adherent air bubbles smoothen samples with rough surface.
- Take care that no water drips onto the upper sample dish when weighing with the help of tweezers.
- In order to reduce the surface tension of water and the friction of the liquid on the wire, add three drops of a common detergent (washing-up liquid) to the measuring liquid (density modification of distilled water occurring due to the addition of tensides can be ignored).
- Oval samples can be held more easily with pincers when you cut grooves into them.
- The density of porous solids may only be determined approximately. Buoyancy errors occur when not all the air is eliminated from the pores during immersion in the measuring fluid.
- To avoid great vibrations of the balance, place sample carefully.
- Avoid static charge, e. g. dry sinker with cotton cloth only.
- If the density of your solid body just deviates slightly from that of distilled water, ethanol may be used as measuring liquid. However, check beforehand whether the sample is solvent-proof. In addition you must observe the applicable safety regulations when working with ethanol.
- To avoid damage by corrosion, don't leave the immersion basket set immersed in liquid for a longer time.