



## **DECLARATION OF CONFORMITY**

**Product Identification:** 

Fluke 179 - Digital Multimeter Family

Including Models: 77 IV, 175, 177

First declared: 2010

## Statement of Conformity:

This Declaration of Conformity is issued under the sole responsibility of the manufacturer, based on sample product type-test, using ISO 9001:2015 Quality Management System. The manufacturer hereby declares this product conforms to the following EU Directives:

Directive 2011/65/EU + 2015/863/EU, UK Regulation S.I. 2012:3032 - Reduction of Hazardous Substances (RoHS) Directive 2014/30/EU, UK Regulation S.I. 2016:1091 - Electromagnetic Compatibility (EMC) Directive 2014/35/EU, UK Regulation S.I. 2016:1101 - Low Voltage (LVD)

## Standards Used:

EN/BS 63000:2018 Assessment of electrical and electronic products for restriction of hazardous substances EN/BS 61326-1:2013 Electrical equipment for measurement, control and laboratory-EMC; General EN/BS 61326-2-2:2013 Particular EMC requirements for portable test, measuring and monitoring equipment EN/BS 61010-1:2010+A1:2019 Safety requirements for measurement, control and laboratory equipment; General EN/BS IEC 61010-2-033:2021/A11:2021 Particular requirements for hand-held Multimeters for measuring Mains voltage

Special Conditions:

None

EU Importer/Distributor

Fluke Europe B.V. PO Box 1186 5602 BD Eindhoven The Netherlands UK Importer/Distributor

Fluke Precision Measurement Ltd. 52 Hurricane Way Norwich, NR6 6JB United Kingdom <u>Manufacturer</u>

Fluke Corporation 6920 Seaway Boulevard M/S 143F Everett, WA 98203; USA

Henry Kuster

EMEA Logistics and Compliance Manager Fluke Europe B.V.

Thomas M. Smith P.E.

Thomas L.

Product Compliance Technical Manager Fluke Corporation