



## EU DECLARATION OF CONFORMITY

Product Identification: Fluke 789, 787B - Process Meter First declared: 2008

Statement of Conformity:

This Declaration of Conformity is issued under the sole responsibility of the manufacturer, based on sample product type-test, using ISO 9001:2015 Quality Management System. The manufacturer hereby declares this product conforms to the following EU Directives:

Directive 2011/65/EU, RoHS Directive 2014/30/EU, Electromagnetic Compatibility (EMC) Directive 2014/35/EU, Low Voltage (LVD)

Standards Used:

EN 50581: 2012 Restriction of the use of hazardous substances EN 61326-1: 2013 Electrical equipment for measurement, control and laboratory-EMC; General EN 61326-2-2: 2013 Particular EMC requirements for portable test, measuring and monitoring equipment EN 61010-1: 2010 Safety requirements for measurement, control and laboratory; General EN 61010-2-030: 2010 Particular requirements for testing and measuring circuits

<u>Special Conditions:</u> None

Importer/Distributor Fluke Europe B.V. Science Park Eindhoven 5110

5692 EC SON

The Netherlands

Thomas & South

Manufacturer

Fluke Corporation 6920 Seaway Boulevard M/S 143F Everett, WA 98203; USA

Thomas M. Smith P.E. Product Compliance Manager, Fluke Corporation, Everett Washington, USA DC20200818083659 - This declaration supersedes all previous declarations for this product. 18 August 2020