

**FLUKE**<sup>®</sup>

## EU DECLARATION OF CONFORMITY

Product Identification:

First declared: 2012

**Fluke 3000FC - Meter Family**

Including Models: 3000, v3000, i3000, a3001, a3002, t3000

Statement of Conformity:

This Declaration of Conformity is issued under the sole responsibility of the manufacturer, based on sample product type-test, using ISO9001:2008 Quality Management System. The manufacturer hereby declares this product conforms to the following EU Directives:

**Directive 2014/30/EU, Electromagnetic Compatibility (EMC)**

**Directive 2014/35/EU, Low Voltage (LVD)**

**Directive 1999/5/EC, Radio and Telecommunications Terminal Equipment (R&TTE)**

Standards Used:

EN 61326-1: 2013 Electrical equipment for measurement, control and laboratory-EMC; General  
EN 61010-1: 2010 Safety requirements for measurement, control and laboratory; General  
EN 61010-2-030: 2010 Particular requirements for testing and measuring circuits  
EN 61010-2-032: 2012 Particular requirements for hand-held and hand manipulated current sensors  
EN 300 328 V1.9.1 Data transmission operation in 2.4GHz ISM band  
EN 301 489-1 V2.1.1 Radio equipment and services  
EN 301 489-17 V3.1.0 Broadband Data transmission systems

Special Conditions:

None

Manufacturer

Fluke  
6920 Seaway Boulevard M/S 143F  
Everett, WA 98203; USA



DC20170310103404

Thomas M. Smith P.E.

Product Compliance Manager, Fluke Corporation

DC20170310103404 - This declaration supersedes all previous declarations for this product.

10 March 2017