



Declaration of Conformity

Date: 3/14/2013

We...

Tektronix Inc.
14200 SW Karl Braun Drive
PO Box 500
Beaverton, Oregon 97007
United States

... declare under sole responsibility that the

Current Probe

TCP305A, P6021A, TCP312A, TCP0030A

... including all options of the above product(s) meets the following requirements unless otherwise noted.

European Union

Low Voltage Directive 2006/95/EC

EN 61010-1:2001 - Safety requirements for electrical equipment for measurement, control and laboratory use. Part 1: General requirements.

EN 61010-2-032:2002 - Electrical equipment for measurement, control, and laboratory use. Part 2-032: Particular requirements for hand-held current clamps for electrical measurement and test.

Conditions and Notes

Mike Flaherty, General Manager, Time Domain Business Unit, Tektronix, Inc.

Martin Baggs, Compliance Manager, Tektronix, Inc.

Send compliance inquiries to:

Tektronix, Inc.	European Union	Australia / New Zealand
Product Compliance Engineering	Tektronix UK Ltd	Baker & McKenzie
PO Box 500, MS/19-045	Western Peninsula, Western Road	Level 27, AMP Centre, 50 Bridge Street
Beaverton, OR 97007 USA	Bracknell RG12 1RF, United Kingdom	Sydney NSW 2000, Australia