

# EC Declaration of Conformity

In accordance with EN ISO 17050-1:2010

We Trumeter Technologies Ltd  
of Trumeter House  
Hornby Street  
Bury  
Lancs  
England  
BL9 5BW

*in accordance with the following Directives:*

2006/95/EC	The Low Voltage Directive (LVD)
2004/108/EC	The Electromagnetic Compatibility Directive (EMC)
2011/65/EU	The Restriction of Hazardous Substances Directive (RoHS 2)

*hereby declare under our sole responsibility that:*

Equipment	APM Panel Meter
Model number(s)	APM-VOLT-APO; APM-VOLT-ANO; APM-VOLT-APN; APM-VOLT-ANN; APM-APM-APO; APM-AMP-ANO; APM-AMP-APN; APM-AMP-ANN; APM-FREQ-APO; APM-FREQ-ANO; APM-FREQ-APN; APM-FREQ-ANN; APM-PROC-ANO; APM-PROC-APO

*are in conformity with the applicable requirements of the following documents:*

<u>Reference</u>	<u>Title</u>	<u>Edition (Date)</u>
IEC 61010-1	Safety requirements for electrical equipment for measurement, control, and laboratory use	3.0 (2010-06)
IEC 61326-1	EMC requirements for equipment designed for measurement, control and laboratory use	2.0 (2012-07)
IEC 61000-4-2	Electrostatic discharge (ESD) immunity test	2.0 (2008-12)
IEC 61000-4-3	Radiated, radio-frequency, electromagnetic field immunity test	3.2 (2010-04)
IEC 61000-4-4	Electrical fast transient/burst immunity test	3.0 (2012-04)
IEC 61000-4-5	Surge immunity test	3.0 (2014-05)
IEC 61000-4-6	Immunity to conducted disturbances, induced by radio-frequency fields	4.0 (2013-10)
IEC 61000-4-8	Power frequency magnetic field immunity tests	2.0 (2009-09)
IEC 61000-4-11	Immunity to voltage dips, interruptions and variations.	2.0 (2004-03)
CISPR-11	Conducted and Radiated Emissions	5.1 (2010-05)

I hereby declare that the equipment named above has been designed to comply with the relevant sections of the above referenced specifications. The unit complies with all applicable Essential Requirements of the Directives.

Signed by: .....

Name: Jeff Allen  
Position: Technical Director  
Done at Bury  
On 30<sup>th</sup> September 2014



Document number: DOC5600  
(issue 2)

<b>Trumeter Technologies Ltd</b>	tel	+44 (0)161 705 4317
Trumeter House	fax	+44 (0)161 705 4319
Europa Business Park		
Barcroft Street	email	info@trumeter.com
Bury BL9 5BT	web	www.trumeter.com

Registration number: 7115948