



## EU DECLARATION OF CONFORMITY

Product Identification:

First declared: 1998

**Fluke 717 - Pressure Calibrator Family**

Including Models: 1G, 15G, 30G, 100G, 300G, 500G, 1000G, 1500G, 3000G, 5000G, 10,000G

Statement of Conformity:

This Declaration of Conformity is issued under the sole responsibility of the manufacturer, based on sample product type-test, using ISO9001:2008 Quality Management System. The manufacturer hereby declares this product conforms to the following EU Directives:

**Directive 2011/65/EU, RoHS**

**Directive 2014/30/EU, Electromagnetic Compatibility (EMC)**

**Directive 2014/35/EU, Low Voltage (LVD)**

Standards Used:

EN 50581: 2012 Restriction of the use of hazardous substances

EN 61326-1: 2013 Electrical equipment for measurement, control and laboratory-EMC; General

EN 61326-2-2: 2013 Particular EMC requirements for portable test, measuring and monitoring equipment

EN 61010-1: 2010 Safety requirements for measurement, control and laboratory; General

EN 61010-2-030: 2010 Particular requirements for testing and measuring circuits

-----

Special Conditions:

None

Manufacturer

Fluke  
6920 Seaway Boulevard M/S 143F  
Everett, WA 98203; USA

DC20170815081906

Thomas M. Smith P.E.

Product Compliance Manager, Fluke Corporation

DC20170815081906 - This declaration supersedes all previous declarations for this product.

15 August 2017