



## EU DECLARATION OF CONFORMITY

Product Identification:

**Fluke 374FC, 375FC and 376FC - True RMS Clamp Meter**

First declared: 2015

Statement of Conformity:

This Declaration of Conformity is issued under the sole responsibility of the manufacturer, based on sample product type-test, using ISO9001:2008 Quality Management System. The manufacturer hereby declares this product conforms to the following EU Directives:

**Directive 2011/65/EU, RoHS**

**Directive 2014/53/EU, Radio Equipment (RED)**

Standards Used:

EN 50581: 2012 Restriction of the use of hazardous substances  
EN 61326-1: 2013 Electrical equipment for measurement, control and laboratory-EMC; General  
EN 61326-2-2: 2013 Particular EMC requirements for portable test, measuring and monitoring equipment  
EN 61010-1: 2010 Safety requirements for measurement, control and laboratory; General  
EN 61010-2-032: 2012 Particular requirements for hand-held and hand manipulated current sensors  
EN 61010-2-033: 2012 Particular requirements for Hand-Held Multimeters for measuring Mains voltage  
EN 300 328 V1.9.1 Data transmission operation in 2.4GHz ISM band  
EN 301 489-1 V1.9.2 Radio equipment and services  
EN 301 489-17 V2.1.1 Broadband Data transmission systems  
-----


Special Conditions:

Products listed above comply with RED harmonized standards EN 300 328 V2.1.1, EN 301 489-1 V2.1.0, and EN 301 489-17 V3.1.1 by inspection. Official testing to confirm in process. Additional standards listed above represent harmonized standards demonstrating compliance to the 2014/30/EU (EMCD) the 2014/35/EU (LVD) directives. These directives are superseded by the Radio Equipment Directive (RED) but the RED has no corollary harmonized standards at this time.

Manufacturer

Fluke  
6920 Seaway Boulevard M/S 143F  
Everett, WA 98203; USA

DC20170619103003

  
Thomas M. Smith P.E.

Product Compliance Manager, Fluke Corporation

DC20170619103003 - This declaration supersedes all previous declarations for this product.

19 June 2017