

**FLUKE®**

EU DECLARATION OF CONFORMITY

Product Identification:

Fluke 902FC - HVAC Clamp Meter

First declared: **2015**

Statement of Conformity:

This Declaration of Conformity is issued under the sole responsibility of the manufacturer, based on sample product type-test, using ISO9001:2008 Quality Management System. The manufacturer hereby declares this product conforms to the following EU Directives:

Directive 2011/65/EU, RoHS

Directive 2014/53/EU, Radio Equipment (RED)

Standards Used:

EN 50581: 2012 Restriction of the use of hazardous substances
EN 61326-1: 2013 Electrical equipment for measurement, control and laboratory-EMC; General
EN 61326-2-2: 2013 Particular EMC requirements for portable test, measuring and monitoring equipment
EN 61010-1: 2010 Safety requirements for measurement, control and laboratory; General
EN 61010-2-032: 2012 Particular requirements for hand-held and hand manipulated current sensors
EN 61010-2-033: 2012 Particular requirements for Hand-Held Multimeters for measuring Mains voltage
EN 300 328 V1.9.1 Data transmission operation in 2.4GHz ISM band
EN 301 489-1 V1.9.2 Radio equipment and services
EN 301 489-17 V2.1.1 Broadband Data transmission systems

Special Conditions:

Products listed above comply with RED harmonized standards EN 300 328 V2.1.1, EN 301 489-1 V2.1.0, and EN 301 489-17 V3.1.1 by inspection. Official testing to confirm in process. Additional standards listed above represent harmonized standards demonstrating compliance to the 2014/30/EU (EMCD) the 2014/35/EU (LVD) directives. These directives are superseded by the Radio Equipment Directive (RED) but the RED has no corollary harmonized standards at this time.

Manufacturer

Fluke
6920 Seaway Boulevard M/S 143F
Everett, WA 98203; USA

DC20170619101444

Thomas M. Smith P.E.

Product Compliance Manager, Fluke Corporation

DC20170619101444 - This declaration supersedes all previous declarations for this product.

19 June 2017