Voltcraft Data Logger

General Information

Device Information

Model:	DL-220THP	Serial ID:	DL220_1609040024
Owner:		Logger ID:	1000

Data Logger Setting

Record Start Condition:	Immediately until memory full	Start Time:	2016-10-06 06:12:11 PM
Sampling Rate:	1Min,0Sec	Stop Time:	2016-10-06 06:21:11 PM
Data Count:	10	Recorded Time:	9Min,0Sec

Alarm Status

Alarm:	Enable					
Temperature:	✗	Low Alarm	30.0°C	✓	High Alarm	70.0°C
Relative Humidity:	✓	Low Alarm	50.0%RH	✓	High Alarm	70.0%RH
Air Pressure:	✓	Low Alarm	1000hPa	✓	High Alarm	1100hpa

Graph

		Minimum	Maximum
Temperature:		25.3°C	25.7°C
Relative Humidity:		59.0%RH	59.9%RH
Air Pressure:		1000hpa	1000hpa



Signature