

**FLUKE®**

EU DECLARATION OF CONFORMITY

Product Identification:

Fluke 368FC - AC leakage Current Clamp Meter

First declared: **2016**

Statement of Conformity:

This Declaration of Conformity is issued under the sole responsibility of the manufacturer, based on sample product type-test, using ISO9001:2008 Quality Management System. The manufacturer hereby declares this product conforms to the following EU Directives:

Directive 2011/65/EU, RoHS

Directive 2014/30/EU, Electromagnetic Compatibility (EMC)

Directive 2014/35/EU, Low Voltage (LVD)

Directive 2014/53/EU, Radio Equipment (RED)

Standards Used:

EN 50581: 2012 Restriction of the use of hazardous substances
EN 61326-1: 2013 Electrical equipment for measurement, control and laboratory-EMC; General
EN 61326-2-2: 2013 Particular EMC requirements for portable test, measuring and monitoring equipment
EN 61010-1: 2010 Safety requirements for measurement, control and laboratory; General
EN 61010-2-032: 2012 Particular requirements for hand-held and hand manipulated current sensors
EN 61557-1: 2007 Electrical safety in low voltage systems; General
EN 61557-13: 2011 Electrical safety in low voltage systems; Measurement of leakage currents
EN 300 328 V1.9.1 Data transmission operation in 2.4GHz ISM band
EN 301 489-1 V1.9.2 Radio equipment and services
EN 301 489-17 V2.1.1 Broadband Data transmission systems

Special Conditions:

None

Manufacturer

Fluke
6920 Seaway Boulevard M/S 143F
Everett, WA 98203; USA

Thomas M. Smith P.E.

Product Compliance Manager, Fluke Corporation

DC20160512134627 - This declaration supersedes all previous declarations for this product.

12 May 2016