

Operating manual

WBGT index HD32.2



Companies / Brands of GHM

Members of GHM GROUP:

GREISINGER
HONSBURG

Martens

IMTRON

Delta OHM

VAL.CO

www.deltaohm.com

Keep for future reference.

TABLE OF CONTENTS

| | |
|---|-----------|
| 1. GENERAL FEATURES | 3 |
| 2. WORKING PRINCIPLE | 6 |
| 3. USER INTERFACES..... | 8 |
| 3.1 THE DISPLAY | 8 |
| 3.2 THE KEYBOARD | 9 |
| 4. WORKING..... | 11 |
| 4.1.1 <i>The measurement “Unit”</i> | 12 |
| 4.1.2 <i>The immediate printing of the data</i> | 12 |
| 4.1.3 <i>The maximum, minimum and medium values of the detected sizes</i> | 12 |
| 4.1.4 <i>Instrument setting</i> | 13 |
| 4.1.5 <i>Start a memorization section (Logging)</i> | 13 |
| 5. THE MAIN MENU | 14 |
| 5.1 INFO ABOUT MENU | 14 |
| 5.2 LOGGING MENU | 16 |
| 5.2.1 <i>Log Interval</i> | 16 |
| 5.2.2 <i>Self Shut-off mode – The automatic switching off</i> | 17 |
| 5.2.3 <i>Start/stop time – The automatic start</i> | 18 |
| 5.2.4 <i>Cancel auto start</i> | 19 |
| 5.2.5 <i>Log File Manager</i> | 21 |
| 5.3 SERIAL MENU (SERIAL COMMUNICATION)..... | 24 |
| 5.3.1 <i>Baud Rate</i> | 25 |
| 5.3.2 <i>Print Interval</i> | 25 |
| 5.4 RESET..... | 26 |
| 5.5 CONTRAST..... | 26 |
| 6. PROBES AND MEASUREMENTS..... | 27 |
| 6.1 <i>Instruction, accuracy and maintenance of the probes</i> | 31 |
| 7. SERIAL AND USB INTERFACES | 32 |
| 7.1 THE DATA MEMORIZATION AND THE DATA TRANSFER TO A PC..... | 34 |
| 7.1.1 <i>Logging function</i> | 34 |
| 7.1.2 <i>Erase function: cancellation of the memorized data</i> | 34 |
| 7.1.3 <i>Print function</i> | 34 |
| 9. INSTRUMENT SIGNALS AND FAULTS..... | 35 |
| 10. BATTERY SYMBOL AND BATTERY REPLACEMENT – MAIN POWER SUPPLY..... | 36 |
| 10.1 WARNING ABOUT THE BATTERIES USE | 36 |
| 11. INSTRUMENT STORAGE | 37 |
| 12. PRINTING OF THE MEASUREMENT REPORT | 38 |
| 13. NOTES ON WORKING AND OPERATING SAFETY..... | 44 |
| 14. TECHNICAL CHARACTERISTICS..... | 45 |
| 15. ORDERING CODES..... | 48 |
| 15.1 PROBES FOR HD32.2 WBGT INDEX..... | 48 |

1. GENERAL FEATURES

HD32.2 is designed for **WBGT** (Wet Bulb Globe thermometer temperature) index analysis in presence or absence of solar radiation. The instrument is provided with three inputs for probes with SICRAM module: the probes have an electric circuit that communicates with the instrument; the sensor calibration data are saved in its permanent memory.

All the SICRAM probes can be inserted in one of the inputs: they are automatically recognized when you switch the instrument on.

The **main features** of the instrument are:

- **Logging:** acquisition and memorization of data inside the instrument. Memory capacity: **64** different **logging sections**, with the possibility to set the acquisition interval of the samples. You can set the **duration of the memorization** and, with **auto-start** function; it's possible to set the starting and finishing date and the hour of data memorization.
- The **measurement unit** of the temperature visualized sizes: °C, °F, °K.
- The system **date** and **hour**.
- The visualization of the **maximum, minimum and medium** statistic parameters and their cancellation.
- The transfer speed of the data through the serial port RS232.

NOTE: The set acquisition interval is the same for all the probes connected to the instrument.

HD32.2 instrument can contemporarily detect the following sizes:

- Globe thermometer temperature **Tg**.
- Wet bulb with natural ventilation temperature **Tn**.
- Ambient temperature **T**.

On the basis of the detected measurements, HD32.2 can calculate:

- **WBGT(in)** index (Wet Bulb Glob Temperature: wet bulb and globe thermometer temperature) in presence or absence of solar irradiation.
- **WBGT(out)** index (Wet Bulb Glob Temperature: wet bulb and globe thermometer temperature) in presence of solar irradiation.

WBGT

Wet Bulb Globe Temperature Meter

HD32.2



1. Inlets for **SICRAM** probes.
2. Inlet for power supply.
3. Back - enlightened graphic display.
4. Navigation key ▲: it allows the navigation through the menus. During the normal working it selects the reset of statistic data.
5. **ESC** key: it allows exiting from the menu or, in case of a submenu, exiting from the current level display.
6. Navigation key ◀: it allows the navigation through the menus. During a normal visualization, it allows to visualize the statistic data: maximum, minimum and medium.
7. **MEM** key: it starts and ends the recording of the data (logging).
8. Navigation key ▼: it allows the navigation through the menus. During the normal working it cancels the reset choice of statistic data.
9. **MENU** key: It allows entering or exiting from the instrument working parameters setting menu.
10. Navigation key ▶: it allows the navigation through the menus.
11. **ENTER** key: it confirms the inserted data inside the menu. During the normal visualization, it allows resetting the statistic data and immediately printing the data on HD40.1 printer.
12. **ON/OFF** key: it switches the instrument on and off.
13. **RS232** and **USB** serial ports.

2. WORKING PRINCIPLE

Microclimate term means the environmental parameters that influence the thermal exchanges between the being and the environments inside limited places and that determinate the so-called “thermal well-being”.

The micro-environmental weather factors together with your own job influence a series of your biologic responses connected to well-being situations (Comfort) or thermal uneasiness (Discomfort).

In fact, the human body tries to keep the thermal balance in equilibrium conditions in order to keep the body temperature on optimal values.

HD32.2 detects the following sizes:

- t_{nv} : wet bulb with natural ventilation temperature probe
- t_g : globe thermometer temperature
- t_a : ambient temperature

It realizes direct measurements with the probes connected to the instrument and it also calculates and directly visualizes **WBGT index**.

WBGT (Wet Bulb Globe Temperature) - wet bulb and globe thermometer temperature - is one of the indexes used for determining the thermal stress to which is submitted a being in a warm environment. It represents the value, with reference to the metabolic waste connected to a special job, over which the being is in a thermal stress condition. WBGT index combines the temperature measurement of natural ventilation wet bulb t_{nw} with the Glob thermometer t_g and, in some situations, with the air temperature t_a . The formula for the calculation is the following one:

- inside and outside the buildings in absence of solar irradiation:

$$WBGT_{close\ environments} = 0,7 t_{nw} + 0,3 t_g$$

- outside the building in presence of solar irradiation:

$$WBGT_{external\ environments} = 0,7 t_{nw} + 0,2 t_g + 0,1 t_a$$

where:

- t_{nw} = natural ventilation wet bulb temperature
- t_g = globe thermometer temperature;
- t_a = air temperature.

The detected data must be compared with the limit values established by the regulation; if they are over passed, it's necessary to:

- directly decrease the thermal stress in the considered working place;
- go on with a detailed analysis of the thermal stress.

In the following table are indicated the limit values of the WBGT thermal stress index taken from ISO 7243 regulation:

| LEVEL CLASS METABOLIC | METABOLIC LEVEL, M | | WBGT LIMIT VALUE | | | |
|-----------------------|---|---|----------------------------------|-----------------------|---------------------------------------|-----------------------|
| | RELATIVE TO A SINGLE SKIN SURFACE W/m ² | TOTAL (FOR A MEDIUM AREA OF 1,8 m ² SKIN SURFACE) W | ACCLIMATIZED BEING TO HEAT °C | | NON- ACCLIMATIZED BEING TO HEAT °C | |
| 0 (RESTING) | M ≤ 65 | M ≤ 117 | 33 | | 32 | |
| 1 | 65 < M ≤ 130 | 117 < M ≤ 234 | 30 | | 29 | |
| 2 | 130 < M ≤ 200 | 234 < M ≤ 360 | 28 | | 26 | |
| 3 | 200 < M ≤ 260 | 360 < M ≤ 468 | STAGNANT AIR 25 | NO STAGNANT AIR 26 | STAGNANT AIR 22 | NO STAGNANT AIR 23 |
| 4 | M > 260 | M > 468 | 23 | 25 | 18 | 20 |

NOTE: THE VALUES ARE STABILIZED, CONSIDERING A MAXIMUM RECTAL TEMPERATURE OF 38° C FOR THE CONSIDERED BEINGS.

To calculate WBGT index, it's necessary that to the instrument are connected:

- **HP3201.2, HP3201 or TP3204S temperature probe with natural ventilation wet bulb.**
- **TP3276.2 or TP3275 globe thermometer.**
- **TP3207.2 or TP3207 wet bulb temperature probe if the detection is made in presence of solar irradiation.**

To measure WBGT index consider the following regulations:

- **ISO 7726**
- **ISO 7243**

3. USER INTERFACES

The user interface is composed of an **LCD graphic back - enlighten display** and start and set keys of the instrument. With a battery power supplying without pressing any key, the back-lighting switches off after 1 minute. To activate it again, press one key. With external power supply, the back-lighting is always activated.

To switch the instrument on and off, press **ON/OFF** key. When you switch it on, the logo and the model of the instrument will be visualize, for some seconds, then you can go to the main visualization.

3.1 THE DISPLAY

| | |
|------------|------------|
| ■■■■ | WBGT Index |
| 2008/11/28 | 08:00:00 |
| Log 00 | 00:00:00 |
| Tn | 15.6 °C |
| Tg | 20.2 °C |
| T | 20.2 °C |
| WBGT (in) | 17.0 °C |
| WBGT (out) | 17.0 °C |

The first line visualizes the **charge condition of the battery**, the **second line indicates the current date and the current hour**. If logging function is activated, the third line indicates the current logging number and the spent time from the logging start.

The **detected sizes** are:

Tn: wet bulb with natural ventilation temperature

Tg: globe thermometer temperature


T: ambient temperature

WBGT (in): WBGT index in absence of direct solar irradiation

WBGT (out): WBGT index in presence of direct solar irradiation

3.2 THE KEYBOARD

The instrument keys have the following functions:



ON-OFF / AUTO-OFF key

ON-OFF: It allows switching the instrument on and off.

When you switch the instrument on, the first screen will be visualized and, after some seconds, the detected measurements will be visualized.

AUTO-OFF: The instrument switches off after 8 minutes from the switching on. When you switch the instrument on, the AUTO-OFF function can be disabled pressing together ESC and ON/OFF keys.



MENU key

It allows entering and exiting from the setting menu of the instrument working parameters.



ENTER key

Inside the menu for confirming the inserted data.

During a normal working:

- it confirms the choice to reset the statistic data.
- it prints the immediate data on HD40.1 printer.



ESC key

You exit from the menu or, if there is a submenu, you exit from the current level visualization.



MEM key

It allows starting and stopping a “logging” section (data memorization); the interval for sending data must be set from the menu.



◀/FUNC key

◀ It allows the navigation through the menus.

FUNC: during a normal visualization, it allows to select some statistic data: maximum, minimum and medium.



▲key

It allows the navigation through the menus. During a normal working, it selects the reset of statistic data.



▼ key

It allows the navigation through the menus. During a normal working, it cancels the choice to reset the statistic data.



▶/UNIT key

▶ It allows the navigation through the menus.

UNIT: it allows selecting the temperature measurement unit: °C, °F, °K.

4. WORKING


Before switching the instrument on, connect the SICRAM probes to the inlets: 8 pin male connector DIN 45326 that is on the bottom of the instrument.

NOTE: The probes must be connected to the instrument off. If you connect a new probe to the instrument already switched on, it isn't recognized, so it's necessary to switch the instrument off and switch it on again.

If you connect a probe with the instrument switched on, you have an acoustic alarm (a beep per second) and you can visualize it on the display in correspondence of a disconnected physical size, the "LOST" message will be visualized.

If you insert more probes of the same type, it will be only consider the first recognized probe: the scansion of the probes, for their recognition, happens from the inlet 1 up the inlet 3.

When you switch the instrument on, on the display appears the following writing for at least 10 seconds:



In addition to Delta Ohm logo are indicated the **instrument code and the firmware version.**

After connecting the probes, switch the instrument on: the display, after 10 seconds, will appear in the measurement visualization mode:

| | |
|---------------------|------------|
| ■■■■ | WBGT Index |
| 2008/11/28 08:00:00 | |
| Log 00 | 00:00:00 |
| Tn | 15.6 °C |
| Tg | 20.2 °C |
| T | 20.2 °C |
| WBGT (in) | 17.0 °C |
| WBGT (out) | 17.0 °C |

Tn: wet bulb temperature with natural ventilation

Tg: globe thermometer temperature, detected by the globe thermometer probe

T: ambient temperature, detected by Pt100 probe

WBGT (in): WBGT index calculated in absence of solar irradiation

WBGT (out): WBGT index calculated in presence of solar irradiation

4.1.1 The measurement “Unit”

Pressing ►/UNIT key, it’s possible to visualize the temperature in degrees °C (Celsius), °F (Fahrenheit) or °K (Kelvin).

4.1.2 The immediate printing of the data

Pressing Enter key, it’s possible to print the immediate data on HD40.1 printer.

Example of immediate printing of the data, obtained using HD40.1 printer.

| | NOTES |
|--------------------------|---|
| ===== | |
| ISO 7243 WBGT Index | Reference regulation |
| ===== | |
| Model HD32.2 WBGT Index | Instrument model |
| Firm.Ver.=01.00 | Instrument firmware version |
| Firm.Date=2008/12/05 | Instrument firmware date |
| SN=12345678 | Instrument serial number |
| ID=0000000000000000 | Identification code |
| ----- | |
| Probe ch.1 description | Description of the probe connected to the inlet 1 |
| Type: Pt100 | |
| Data cal.:2008/10/01 | |
| Serial N.:08109450 | |
| ----- | |
| Probe ch.2 description | Description of the probe connected to the inlet 2 |
| Type: Pt100 Tg 50 | |
| Cal. date.:2008/10/01 | |
| Serial N.:08109452 | |
| ----- | |
| Probe ch.3 description | Description of the probe connected to the inlet 3 |
| Type: Pt100 Tw | |
| Cal. date.:2008/10/01 | |
| Serial N.:08109454 | |
| ----- | |
| Date=2008/11/21 15:00:00 | Date and hour |
| Tnw 21.2 °C | Wet bulb temperature with natural ventilation |
| | Globe thermometer temperature |
| Tg 24.9 °C | Wet bulb temperature |
| Ta 31.3 °C | WBGT in absence of direct solar irradiation |
| WBGT (i) 22.3 °C | WBGT in presence of direct solar irradiation |
| WBGT (o) 23.0 °C | |
| ===== | |
| Notes: | |
| ===== | |

4.1.3 The maximum, minimum and medium values of the detected sizes

Pressing ◀/FUNC key, it’s possible to visualize the maximum, minimum and medium value of the detected sizes.

To reset the statistic data, press ◀/FUNC key till “Clear Func? Yes No” writing doesn’t appear.

Select Yes with ▲ ▼ keys and confirm with ENTER key.

NOTE: Once selected, for example, *max*, all the visualized sizes indicate the maximum value. **The average is calculated on the number of samples belonging to the first five minutes and then, considering the current average.**

4.1.4 Instrument setting

To set the instrument up, you enter to the main menu, pressing **MENU** key. For more details, see the chapter 5.

4.1.5 Start a memorization section (Logging)

To start a **Logging** section, press **MEM** key: the key starts and stops the memorization (Logging) of a data block that will be kept into the instrument internal memory. The frequency with which the data are memorized is set with "**Log interval**" menu parameter. The memorized data between a start and a following stop represent a measurement block.

With the memorization function on, **LOG and number of logging section** appears on the display; a beep is given out during each memorization.

To finish the logging, press **MEM** key again.

The instrument can switch off during the logging between an acquisition and the following one: the function is controlled by **Auto_shut_off_Mode** parameter. With a memorization interval less than one minute, the instrument remains always on during the logging; with an interval of at least one minute, it switches off between an acquisition and the following one.

5. THE MAIN MENU

To visualize the programming menu, press **MENU** key:

```
■■■■ WBGT Index
2008/11/10 08:00:00
MAIN MENU
Info
Logging
Serial
Reset
Contrast
<ESC> exit/cancel
```

If you don't press any key for at least 2 minutes, the instrument will be back to the main visualization.

To select a heading, use ▲▼ arrow keys and press **ENTER**.

To exit from the selected heading and be back to the previous menu level, press **ESC**.

To exit directly from the main menu, press **MENU** again.

5.1 INFO ABOUT MENU

Pressing **MENU** key, you enter into the main menu. To enter to Info menu, select **Info** heading with ▲▼ keys and press **ENTER**.

```
■■■■ WBGT Index
2008/11/10 08:00:00
INFO
Info Instrument
Info Probe
Time/Date
<UP> <DOWN> select
<ENTER> confirm
<ESC> exit/cancel
```

Selecting **Info Instrument**, you can visualize the information involving the instrument: instrument code and operative program, firmware version and date, serial number, instrument calibration date and identification code.

```
■■■■ WBGT Index
2008/11/10 08:00:00
INFO INSTRUMENT
Model HD32.2
Firm.Ver.=01.00
Firm.Date=2008/06/30
Ser. Number=08010000
Calib: 2008/11/10
ID: 0000000000000000
```

To modify the **ID**, press **ENTER**. Select the heading that you want to change with ◀▶ arrows and modify it with ▲▼ arrows. Go on with the other headings and, at the end, confirm with **ENTER** key.

Selecting **Info Probe** you can visualize the information involving the probes connected to the inlets:

```
INFO PROBE
Ch.1:Pt100Tg_50
Cal=2008/06/20
SN=08018422
Ch.2:Pt100
Cal=2008/06/21
SN=08018423
Ch.3:Pt100Tw
Cal=2008/06/20
SN=08018424
```

INFO ABOUT PROBE:

Description of the probe connected to the inlet 1.
Calibration data of the probe connected to the inlet 1.
Serial number of the probe connected to the inlet 1.

Description of the probe connected to the inlet 2.
Calibration data of the probe connected to the inlet 2.
Serial number of the probe connected to the inlet 2.

Description of the probe connected to the inlet 3.
Calibration data of the probe connected to the inlet 3.
Serial number of the probe connected to the inlet 3.

To go back to the main menu, press **ESC**. To exit from the menu, press **MENU**.

Time/Date allows to set the date and the hour that will appear on the upper part of the display.

To enter into **Time/date**, go on as indicated below:

1. using ▼▲ arrow keys, select **Time/date** heading
2. press **ENTER**
3. the message will be visualized

```
■■■■
2008/11/10 08:00:00
enter date/time
<- arrows change ->
<ENTER> confirm
and set 00 seconds !
year/mm/dd hh:mm
2008/11/28 11:10:26
```

4. use ◀▶ arrows to select the data to set (year/month/day – hour : minutes)
5. once selected, the data will start flashing;
6. using ▼▲ arrows, insert the right value;

7. press **ENTER** to confirm and go back to the main menu;
8. or press **ESC** to go back to the menu, without modifying anything;
9. press **MENU** to exit directly from the main menu.

NOTE: With reference to the time, you can set hour and minuted, the seconds are always set at 00 (set 00 seconds!).

5.2 LOGGING MENU

Press **MENU** key to enter into the main menu;

- Select **Logging** heading using **▲ ▼** keys;
- press **ENTER**: the submenu for the parameters setting involving the Logging section (to aquired) is visualized.

```

■■■■ WBGT Index
2008/11/10 08:00:00
LOGGING MENU
Log interval
Self shut_off mode
Start/stop time
Cancel auto start
Log file manager
<ESC> exit/cancel

```

5.2.1 Log Interval

Through this heading, it's possible to set the LOG interval (interval between two acquisition following samples): to set, go on as indicated below:
 enter into **LOGGING** submenu (previous paragraph), select **Log Interval** heading using **▲ ▼** arrow keys:

```

■■■■ WBGT Index
2008/11/10 08:00:00
LOGGING MENU
input LOG interval
as h:mm:ss (1h max)
arrows to correct
or <ESC> now set at:
0:00:15

```

1. Using **▲ ▼** arrow keys select the interval duration that is between a minimum of 15 seconds and a maximum of one hour.
2. Press **ENTER** to confirm and go back to Logging menu.
3. To go back to **Logging** menu without modifying anything, press **ESC**.
4. To go back to the main menu, press **ESC** again.
5. To exit directly from the menu, press **MENU**.

The values that you can set are the following ones: 15 seconds - 30 seconds - 1 minute - 2 minutes - 5 minutes - 10 minutes - 15 minutes - 20 minutes - 30 minutes - 1 hour

| Memorization interval | Memory capacity | Memorization interval | Memory capacity |
|-----------------------|-----------------------------|-----------------------|----------------------------|
| 15 seconds | About 11 days and 17 hours | 10 minutes | About 1 year and 104 days |
| 30 seconds | About 23 days and 11 hours | 15 minutes | About 1 year and 339 days |
| 1 minute | About 46 days and 22 hours | 20 minutes | About 2 years and 208 days |
| 2 minutes | About 93 days and 21 hours | 30 minutes | About 3 years and 313 days |
| 5 minutes | About 234 days and 17 hours | 1 hour | About 7 years and 261 days |

5.2.2 Self Shut-off mode – The automatic switching off

Self shut-off mode heading controls the instrument automatic switching off mode during the logging between the acquisition of a sample and the following one. **With an interval less than 60 seconds, the instrument will always remain on.** With intervals upper or equal to 60 seconds, it's possible to choose switching the instrument off between the memorizations: it will switch on in correspondence of the sampling time and it will switch off after some seconds, so extending the life of the batteries.

Once you entered into **LOGGING** sub menu (previous paragraph), select **Self shut_off mode** heading using ▲ ▼ arrow keys:

- If the set **Log Interval** (see the previous paragraph) is less than 60 seconds, there will be visualized:

```

■■■■  WBG7 Index
2008/11/10 08:00:00
Log interval<60 sec
During log session
the instrument
will STAY ON
between samples

```

- If the set **Log Interval** (see the previous paragraph) is upper or equal to 60 seconds, there will be visualized:

```

■■■■  WBG7 Index
2008/11/10 08:00:00
Log interval>=60 sec
During log session
the instrument
will SHUT OFF
between samples

```

1. Pressing ▲ ▼ arrows, you can select:

STAY ON (the instrument remains on)
SHUT OFF (the instrument remains off)

2. To go back to **Logging** menu, press **ESC**.
3. To go back to the main menu, press **ESC** again.
4. To exit directly from the menu press **MENU**.

5.2.3 Start/stop time – The automatic start

The start and the end of the memorization can be programmed inserting the date and the hour. The function offers, as start hour, the current hour will increase of 5 minutes: to confirm, press <ENTER>, vice versa set the date and the hour using the arrows. So it's requested to set the data for ending the memorization: in a default mode, the instrument offers the start hour increased of 10 minutes. The values offered in default mode are such to allow the user arranging the instrument for the measurement.

NOTE: in a default mode, the set time is more than 5 minutes with reference to the current hour.

For the setting, go on as indicated below.

Once entered into **LOGGING** submenu, select **Start/Stop time heading using ▲ ▼** arrow keys: "Enter start time" message will be visualized as indicated below:

```

■■■■ WBGT Index
2008/11/10 08:00:00
enter start time
arrows to correct
<ENTER> confirm
default= 5m>RealTime

2008/11/28 10:29:00

```

1. using ◀ ▶ arrow key select the data to modify (year/month/day and hour/minutes/seconds);
2. once selected, the data will start flashing;
3. change the valued with ▼ ▲ keys;
4. confirm pressing **ENTER**;
5. to be back to **Logging** menu without modifying anything, press **ESC**;
6. to be back to the main menu, press **ESC** again;
7. to exit directly from the main menu, press **MENU**.

After setting the time of the memorization start, there will be visualized the request screen for inserting the end memorization time (enter stop time):

```

■■■■ WBGT Index
2008/11/10 08:00:00
enter stop time
arrows to correct
<ENTER> confirm
default=10m>RealTime

2008/11/28 10:39:00

```

1. using ◀ ▶ arrow key select the data to modify (year/month/day and hour/minutes/seconds);
2. once selected, the data will start flashing;
3. change the valued with ▼ ▲ keys;
4. confirm pressing ENTER;
5. to be back to **Logging** menu without modifying anything, press ESC;
6. to be back to the main menu, press ESC again;
7. to exit directly from the main menu, press MENU.

NOTE: in a default mode, the set end acquisition time is more than 10 minutes with reference to the start time of Logging section..

1. Once set both the values, the summary of the times will be visualized: date and hour of start and end of LOG section.

```

■■■■  WBG7 Index
2008/11/10 08:00:00
<ENTER> confirm
Start time
2008/11/28 10:29:00
End time
2008/11/28 10:39:00

<ESC> exit/cancel

```

2. Press ENTER to confirm or ESC to exit without activating the automatic start: in both the cases, you go back to LOGGING menu.
3. Press MENU to exit directly from the main menu.

When the instrument starts a LOG section in an automatic mode, a beep is given out for each acquisition and, in the upper part of the display, LOG flashing heading will appear.

To stop the section before the set stop time, press MEM key.

To cancel the settingf of the automatic start, use **Cancel auto start function, described in the previous paragraph.**

NOTE: the automatic logging section starts also if the instrument is off. If, during the start of automatic logging section, the instrument is off, it switches on some minutes before the start time and, at the end of the logging, it remains on. If it's supplied by a battery, it switches off after some minutes spent in no working conditions at the end of the logging section.

To set the auto-switching, see the paragraph 5.2.2.

5.2.4 Cancel auto start

Once set the start and end times of LOG section, you can avoit the auto start of the section through **Cancel auto start** heading.

Once entered into **LOGGING** submenu:

1. select, using ▲ ▼ arrow keys, **Cancel auto start** heading
2. a message containing the start and end time of LOG section will be visualized:

```
■■■■ WBGT Index
2008/11/10 08:00:00
Self-timer abort
Start scheduled at
2008/11/28 10:29:00
Stop scheduled at
2008/11/28 10:39:00
Press ARROW to
delete schedule
```

3. pressing ▲ key the following message will be visualized: "Self timer not active";

```
■■■■ WBGT Index
2008/11/10 08:00:00

Self timer
not active
```

4. press **ENTER** to cancel the auto start;
5. press **ESC** to exit without cancelling the auto start;
6. press **ESC** again to go out from the different submenus;
7. or press **MENU** to exit directly from the main menu.

After cancelling the auto start time, to set a new one, see the previous paragraph.

5.2.5 Log File Manager

Through this heading, it's possible to manage the acquired log sections: the instrument allows printing the acquired data files (*Print selected log*) and cancelling the whole memory (*Erase ALL logs*).

The instrument can memorize up to 64 sections of LOG numbered progressively from 00 to 63: the list of the sections is placed on 4 lines and 4 columns. If there are more than 16 sections, using **MEM** key you can visualize the following screen. Downwards, at the right, is indicated the current page (0, 1, 2 or 3) and the total number of pages with the memorized data: in the example, "0/3" means that you must print the page 0 on the three pages with memorized data.

```
LOG FILE                                0/3
00 - 01 - 02 - 03
04 - 05 - 06 - 07
08 - 09 - 10 - 11
12 - 13 - 14 - 15
Date:
2008/11/28 08:59:40
rec: 000039
<MEM> to change Page
```

Once you enter into the submenu **LOGGING**:

1. select **Log File manager** heading through ▲▼ arrow keys: the following submenu will be visualized:

```
■■■■ WBGT Index
2008/11/10 08:00:00
LOG FILE MANAGER
Print selected log
Erase ALL logs
Log time
```



HD40.1

2. to select a menu heading, use ▲ ▼ arrow keys;
3. press **ENTER** to confirm;
4. press **ESC** to go back to the menu;
5. press **MENU** to exit immediately from the main menu.

NOTE: you can connect a PC to the serial port RS232C of the instrument or the **HD40.1 printer**. Before starting to print through the port RS232C, you must set the baud rate. To do this, select **Baud Rate** heading from **Serial** menu (see the paragraph 5.3.1 The Baud Rate) and select the maximum value equal to **38400 baud**. For the connection to a printer, use the maximum value tolerated by the printer.

The communication between the instrument and the PC or between the instrument and the printer works only if the baud rate of the instrument and the one belonging to the connected device (computer or printer) are the same.

Example of printing of a logging section, obtained with HD40.1 printer

```

Log number: 1
=====
  ISO 7243 WBGT Index
=====
Model HD32.2 WBGT Index
Firm.Ver.=01.00
Firm.Date=2008/12/05
SN=12345678
ID=0000000000000000
-----
Probe ch.1 description
Type: Pt100
Data cal.:2008/10/01
Serial N.:08109450
-----
Probe ch.2 description
Type: Pt100 Tg 50
Data cal.:2008/10/01
Serial N.:08109452
-----
Probe ch.3 description
Type: Pt100 Tw
Data cal.:2008/10/01
Serial N.:08109454
=====
From=2008/11/21 15:00:00
To =2008/11/21 16:30:00
Tot. rec.= 000360

**      max value      **
Tnw          21.2 °C

Tg           24.9 °C
Ta           31.3 °C
WBGT (i)    22.3 °C
WBGT (o)    23.0 °C

**      min value      **
Tnw          21.0 °C
Tg           24.9 °C
Ta           23.5 °C
WBGT (i)    22.2 °C
WBGT (o)    22.6 °C

**      avg value      **
Tnw          21.1 °C
Tg           24.9 °C
Ta           30.3 °C
WBGT (i)    22.3 °C
WBGT (o)    22.8 °C
=====
Notes:

=====

```

NOTES

```

Number of logging sections

Reference rules

Instrument model
Version of the instrument firmware
Date of the instrument firmware
Instrument serial number
Identification code

Description of the probe connected to the inlet 1

Description of the probe connected to the inlet 2

Description of the probe connected to the inlet 3

Logging start day and hour
Logging end day and hour
Number of samples acquired by the instrument

Maximum value of the acquired data
Maximum temperature of wet bulb with natural ventilation
Maximum temperature of globe thermometer
Maximum temperature of wet bulb
Maximum WBGT in absence of direct solar irradiation
Maximum WBGT in presence of direct solar irradiation

Minimum value of the acquired data
Minimum temperature of wet bulb with natural ventilation
Minimum temperature of globe thermometer
Minimum temperature of wet bulb
Minimum WBGT in absence of direct solar irradiation
Minimum WBGT in presence of direct solar irradiation

Medium value of the acquired data
Medium temperature of wet bulb with natural ventilation
Medium temperature of globe thermometer
Medium temperature of wet bulb
Medium WBGT in absence of direct solar irradiation
Medium WBGT in presence of direct solar irradiation

```

NOTE: The printing of a logging section on HD40.1 printer contains only the statistic data. To visualize all the acquired data, it's necessary to download the data using DeltaLog10 software.

Print selected log

Selecting this heading you can visualize the logging sections that are into the instrument:

```
LOG FILE                                0/3
00 - 01 - 02 - 03
04 - 05 - 06 - 07
08 - 09 - 10 - 11
12 - 13 - 14 - 15
Date:
2008/11/28 08:59:40
rec: 000039

<MEM> to change Page
```

1. select the log to print, using ▲▼◀▶ arrows and MEM key to go to the following page.
2. once selected a file, in the lower part of the display, are indicated the date and the hour of acquisition start and the number of samples contained into the file (Rec). **The files are memorized in a growing order.** Each file is only identified by the date and the hour **indicated on the display.** In the example above, the file 00 is selected: the memorization started at 08:50:40 on 11/28/2008. The file contains 39 samples.
3. the data transfer message is visualized for some seconds, then the instrument go back to **Print selected log** screen to select another log to print.
4. repeat the process to print the sections you are interested in or press ESC to exit from this menu level.
5. press MENU to exit immediately from the main menu.

Erase all memory

Selecting this heading, “ERASE ALL FILES” message will be visualized:

```
■■■■ WBGT Index
2008/11/10 08:00:00
ERASE ALL FILE

<MEM> confirm
<Esc> exit
```

1. press MEM to erase all the files;
2. press ESC to erase the operation and go back to the upper menu level;
3. press MENU to exit immediately from the main menu.

2) Log time (Set time for the memorization)

It represents the duration of the memorization: after the set time, the memorization stops. The memorization can be finished before the expiration of the set time, pressing **MEM** key.

To disable the function, set the time at 0:00:00. In this case, the memorization stops pressing MEM key or when the memory is full.

```
■■■■ WBGT Index
2008/11/10 08:00:00
LOG TIME
as h:mm:ss (1h max)
arrows to correct
or <ESC> now set at:
00:00:00
```

With the arrows modify the set time, the maximum value admitted is 1 hour.

Confirm with **ENTER** key.

Press **ESC** to exit from this menu level without modifying anything.

Press **MENU** to exit immediately from the main menu.

5.3 SERIAL MENU (SERIAL COMMUNICATION)

Serial sub menu allows setting the data transfer speed by serial means (*Baud rate*) and the record printing interval (*Print Interval*).

LOG sections can be downloaded to a PC by **RS232** serial connection or **USB** connection.

In case of serial connection, the transmission speed is set by the user (see the following paragraphs) and, however, can't be more than 38400 bps.

In case of USB connection, the transmission speed is fixed at 460800 bps.

After downloading the data on your PC, by the relative software, the data are elaborated by the software for the graphical visualization and for the calculation of the comfort/stress indexes.

The instrument can be directly connected to **HD40.1 printer**.

To enter into *Serial* submenu, go on as indicated below:

1. Press **MENU** key of the instrument;
2. Select, using ▼ ▲ arrow keys, *Serial* heading;
3. press **ENTER**;
4. *Serial* submenu will be visualized.

```
■■■■ WBGT Index
2008/11/10 08:00:00
SERIAL COM MENU
Baudrate
Print Interval

<UP> <DOWN> select
<ENTER> confirm
<ESC> exit/cancel
```


5.3.1 Baud Rate

Baud Rate represents the speed used for the serial communication with the PC.

To set **Baud rate**, go on as indicated below:

1. select the heading with ▼▲ arrow keys;
2. press **ENTER**: the following message will be visualized:

```
■■■■ WBGT Index
2008/11/10 08:00:00

set Baudrate
arrows to correct
Or <ESC> now set at:
38.400
```

3. set the value through ▼▲ arrow keys;
4. press **ENTER** to confirm and go back to the previous screen or press **ESC** not to modify the value and exit for the menu heading;
5. press **ESC** over and over to exit from the menu different levels;
6. press **MENU** to exit immediately from the main menu.

WARNING

The communication between the instrument and the PC (or serial port printer) works only if the instrument and PC Baud rates are the same. If USB connection is used, this parameter value is automatically set.

NOTE: When setting the baud-rate, check the printer speed capacity.

5.3.2 Print Interval

To set **Print Interval**, go on as indicated below:

1. use ▲▼ arrow keys to select the heading;
2. press **ENTER**: the following message will be visualized;

```
■■■■ WBGT Index
2008/11/10 08:00:00
SERIAL COM MENU
input PRINT interval
as h:mm:ss (1h max)
arrows to correct
or <ESC> now set at:
0:00:00
```

3. using ▼▲ arrow keys, set the value;
4. press **ENTER** to confirm and go back to the previous screen or press **ESC** not to modify the value and exit from the menu heading;
5. press **ESC** over and over to exit from menu different levels;

6. press **MENU** to exit immediately from the main menu.

The print interval can be set from 0 seconds to one hour:

0 s - 15 s - 30 s - 1 min. - 2 min. - 5 min. - 10 min. - 15 min. - 20 min. - 30 min. - 1 hour.

5.4 RESET

To enter into **Reset** submenu in order to carry out a complete reset of the instrument, go on as indicated below:

1. press **MENU** key of the instrument;
2. select **Reset** heading, using **▼ ▲** arrow keys;
3. press **ENTER**: the message will be visualized:

```
■■■■ WBGT Index
2008/11/10 08:00:00

0)Reset

<Up-Down>: select
<Enter>: confirm
```

4. press **▼ ▲** arrow key to select **Reset** heading;
5. press **ENTER** to confirm or press **ESC** over and over to exit from the menu different levels;
6. press **MENU** to exit immediately from the main menu.

5.5 CONTRAST

This menu heading allows increasing or decreasing the contrast on the display:

To enter into **Contrast** submenu, go on as indicated below:

1. Press **MENU** key of the instrument;
2. Use **▲ ▼** arrow keys to select **Contrast.**;
3. Press **ENTER**;
4. The following message will be visualized.

```
■■■■ WBGT Index
2008/11/10 08:00:00
LCD CONTRAST
<- arrows change ->
<ESC> exit/cancel
Contrast Adjust: 012
```

5. use **◀ ▶** arrow key to decrease or increase the contrast;
6. press **ENTER** or **ESC** to go back to the main menu;
7. press **MENU** to exit immediately from the main menu.

6. PROBES AND MEASUREMENTS

Necessary probes for **WBGT** measurement:



**HP3201.2, HP3201
or TP3204S**

TP3276.2 or TP3275

TP3207.2 or TP3207

TP3207.2, TP3207

Temperature probe

Sensor type:

Thin film Pt100

(*) Accuracy:

Class 1/3 DIN

Measurement range:

-40 ÷ 100 °C

Connection:

4 wires plus SICRAM module

Connector:

8-pole female DIN45326

Cable:

Only TP3207 (2m)

Dimensions:

Ø=14 mm L= 150 mm (TP3207.2), L= 140 mm (TP3207)

(**) Response time T₉₅

15 minutes

(*) The probe is calibrated, the calibration data are kept into the SICRAM module memory.

(**) T₉₅ response time is the time spent to reach the 95% of the final value. The measurement of the response time is done with a negligible air speed (motionless air).

TP3276.2, TP3275

Globe thermometer probe $\varnothing=50$ mm (TP3276.2), $\varnothing=150$ mm (TP3275)

Sensor type: Pt100

(*) Accuracy: Class 1/3 DIN

Measurement range: $-10 \div 100$ °C

Connection: 4 wires plus SICRAM module

Connector: 8-pole female DIN45326

Cable: Only TP3275 (2m)

Stem dimension: $\varnothing=8$ mm L= 170 mm (TP3276.2),

$\varnothing=14$ mm L= 110 mm (TP3275)

(**) Response time T_{95} 15 minutes



(*) The probe is calibrated, the calibration data are kept into the SICRAM module memory.

(**) T_{95} response time is the time spent to reach the 95% of the final value. The measurement of the response time is done with a negligible air speed (motionless air).

HP3201.2, HP3201

Wet bulb probe with natural ventilation

- Sensor type: Pt100
- (*) Accuracy: Class A
- Measurement range: 4 °C ÷ 80 °C
- Connection: 4 wires plus SICRAM module
- Connector: 8-pole female DIN45326
- Cable: Only HP3201 (2m)
- Stem dimension: Ø=14 mm L= 170 mm (HP3201.2), L= 110 mm (HP3201)
- Cotton wick length: about 10 cm
- Tank capacity: 15 cc
- Tank autonomy: 96 hours con RH=50%, t = 23°C
- (**) Response time T₉₅ 15 minutes

HP3201.2, HP3201 wet bulb probe with natural ventilation

For the start up go on as indicated below:

- Remove the cover, **the cover isn't screwed.**
- Insert the cotton wick into the temperature probe, the cotton wick was previously dipped with distilled water.
- Fill the case up till $\frac{3}{4}$ with **distilled water.**
- Close the case cover.
- **Warning:** don't turn the probe in the vertical direction because the distilled water can exit.
- The cotton wick must project from the temperature probe for about 20mm.
- During the time, the cotton wick is going to calcify (to become hard) and so it's necessary to replace periodically it.



(*) The probe is calibrated, the calibration data are kept into the SICRAM module memory.

(**) T₉₅ response time is the time spent to reach the 95% of the final value. The measurement of the response time is done with a negligible air speed (motionless air).

TP3204S

Natural ventilation wet bulb probe for long-lasting measurements

Sensor type: Pt100

(*) Accuracy: Class A

Measurement range: 4 °C ÷ 80 °C

Connection: 4 wires plus SICRAM module

Connector: 8-pole female DIN45326

Cable: 2 m

Dimensions: L x W x H=140 x 65 x 178.5 mm (reservoir + bottle)

Cotton wick length: about 10 cm

Tank capacity: 500 cc



Tank autonomy: 15 days @ t = 40 °C

(**) Response time T₉₅ 15 minutes

TP3204S wet bulb probe with natural ventilation

For the start up go on as indicated below:

- Remove the sensor cap (**the cap is not screwed**).
- Insert the cotton wick, previously dipped with distilled water, into the temperature probe. The cotton wick must protrude from the probe for about 20 mm.
- Replace the cap.
- Fill the bottle with 500 cc of **distilled water**.
- Turn the probe over and firmly screw the bottle to the probe reservoir.
- Turn the probe quickly (to avoid water spillage).
- Secure the probe to the **HD32.2.7.1** support by using the two screws at the bottom of the probe.



For measurements in presence of solar irradiation, use the protection shield from solar radiations. The cotton wick calcifies (becomes hard) with time: replace it periodically.

(*) The probe is calibrated, the calibration data are kept into the SICRAM module memory.


(**) T₉₅ response time is the time spent to reach the 95% of the final value. The measurement of the response time is done with a negligible air speed (motionless air).

Measurement

In the place where you want to measure, you must assemble the tripod and the probes necessary for the measurement. You set the instrument up and then you start measuring. If the measurement has to be done in other places, you have to move all the equipment in the new measurement position.

At the end of the measurement or after some times, the acquired data are transferred to the PC for the elaboration and to allow writing the measurement report/s.

6.1 *Instruction, accuracy and maintenance of the probes*

- Don't expose the probe to gas or liquids that could corrode the probe material. After the measurement clean accurately the probes.
 - Don't fold the connector, using force upwards or downwards.
 - Respect the correct probe polarity.
 - During the insertion of the probe's connector into the instrument, don't fold or use force against the contacts.
 - Don't fold the probes and don't distort them or let them fall: they can damage in an irreparable manner.
 - Use a probe suitable to the measurement type that you want to realize.
 - For a reliable measurement, avoid too speed temperature variations.
- 
- Some sensors aren't isolated with referent to the external sheathing, pay attention not to be in contact with items under tension (up to 48V): it could be dangerous for the instrument and for the operator who can be electrocuted.
 - Avoid measuring in presence of high-frequency sources, microwave or great magnetic fields, because they will be not so reliable.
 - The instrument is water-proof, you don't dip it into the water. If it falls into the water, check that there isn't any infiltration.

7. SERIAL AND USB INTERFACES

HD32.2 has a RS-232C serial interface, galvanically isolated and it has a USB 2.0 interface. As an optional, we can supply, under request:

- RS-232C serial connection cable (code **HD2110/RS**) with M12 connector from instrument side and 9-pole female connector sub D from the PC side.
- USB 2.0 connection cable (code **HD2110/USB**) with M12 connector from instrument side and USB 2.0 connector from the PC side.

The USB connection requires the previous installation of a driver included in the DeltaLog10 software package. **Install the driver before connecting the USB cable to the PC** (follow the instructions included in the software package).

The instrument standard RS232 serial transmission parameters are:

- Baud rate 38400 baud
- Parity None
- N. bit 8
- Stop bit 1
- Protocol Xon / Xoff.

It's possible to change the speed of RS232C serial data transmission acting on "*Selection of serial transfer speed (Baud Rate)*" inside the menu (see the menu of the chapter **5.3.1 Baud Rate**). The possible values are: 38400, 19200, 9600, 4800, 2400, 1200. The other transmission parameters are fixed.

The USB 2.0 connection does not require the setting of parameters (Baud rate = 460800 fixed).

The port selection is directly done by the instrument: if USB port is connected to a PC, RS232 serial port is automatically excluded and vice versa.

The instruments are provided with a complete set of commands and data request to send by PC.

All the commands transmitted to the instrument must have the following structure:

XXCR where: **XX is the command code and CR is the Carriage Return (ASCII 0D).**

The command characters XX are only capital, the instrument responses with "&" if the command is correct, with "?" for each characters combination that is wrong.

The response string of the instrument is finished, entering CR (Carriage Return) and LF (Line Feed) commands.

Before entering the commands to the instrument through the serial port, we suggest you to block the keyboard to avoid working conflicts: use P0 command. At the end of the process, restore the keyboard use with P1 command.

| Command | Response | Description |
|---------|-------------------------|---|
| P0 | & | Ping (locks the instrument keyboard for 70 seconds) |
| P1 | & | Unlocks the instrument keyboard |
| S0 | | |
| G0 | Model HD32.2 | Instrument model |
| G1 | M=WBG Index | Model description |
| G2 | SN=12345678 | Instrument serial number |
| G3 | Firm.Ver.=01.00 | Firmware version |
| G4 | Firm.Date=2008/10/12 | Firmware date |
| G5 | cal 2008/11/10 10:30:00 | Calibration date and time |
| C1 | | Probe 1 type, serial number, calibration date |
| C2 | | Probe 2 type, serial number, calibration date |

| Command | Response | Description |
|---------|---|---|
| C3 | | Probe 3 type, serial number, calibration date |
| GC | | Print instrument's heading |
| GB | ID=0000000000000000 | User code (set with T2xxxxxxxxxxxxxxxx) |
| H0 | Tw= 19.5 øC | Print wet bulb temperature |
| H1 | Tg= 22.0 øC | Print globe thermometer temperature |
| H2 | Ta= 21.6 øC | Print air temperature (dry bulb); |
| H7 | WGBT (i)= 23.0 øC | Print indoor WGBT (without solar radiation) |
| H8 | WGBT (o)= 24.0 øC | Print outdoor WGBT (with solar radiation) |
| LN | A00 -A01 -A02 -A03 - .. - .. - .. - .. - .. - .. - .. - .. - .. - .. - .. - .. - | Print instrument memory map: if a section is allocated a number is displayed, if it is free 2 points (..) are displayed. |
| LFn | !Log n.= 0!started on:!2006/01/01 00:37:32 | Print memory n section status. The number, the storage start date and time are displayed. (n= hexadecimal number 0-F). If the section is empty: "-->No Log Data<--" |
| LDn | | Print data stored in section n. If the section is empty: "-->No Log Data<--" |
| LE | & | Cancel stored data |
| K1 | & | Immediate data printing |
| K0 | & | Stop printing data |
| K4 | & | Start logging data |
| K5 | & | Stop logging data |
| KP | & | Auto-power-off function=ENABLE |
| KQ | & | Auto-power-off function=DISABLE |
| WC0 | & | Setting SELF off |
| WC1 | & | Setting SELF on |
| RA | Sample print = 0sec | Reading of PRINT interval set |
| RL | Sample log = 30sec | Reading of LOG interval set |
| WA# | & | Setting PRINT interval. # is a hexadecimal number 0..D that represents the position of the interval in the list 0, 1, 5, 10, ..., 3600 seconds. |
| WL# | & | Setting LOG interval. # is a hexadecimal number 1..D that represents the position of the interval in the list 15, ..., 3600 seconds. |

7.1 THE DATA MEMORIZATION AND THE DATA TRANSFER TO A PC

HD32.2 can be connected to RS232C serial port or USB port of a PC and can exchange data and information through DeltaLog10 software that works with Windows. The values measured can be printed (ENTER key) or stored in the memory of the instrument by using the *Logging* function (MEM key). The memorized data can be transferred to the PC in a second time.

7.1.1 Logging function

Logging function allows memorizing the measurements detected by the probes connected to the inlets. The interval between two following measurements can be set between 15 seconds and 1 hour. The memorization start is obtained pressing **MEM** key; the stop is obtained pressing the same key: so, the memorized data are a continuous block of data.

See the menu headings description in the chapter 5. MEIN MENU".

If the *Self Shut-off* option is activated between two memorizations (see the paragraph 5.22 *Self Shut-off mode*), pressing **MEM** key, the instrument memorizes the first data and the switches off, 15 seconds before the following memorization instant, the instrument switches on to acquired the new sample and then it switches off.

The memorized data can be transferred to the PC (see the paragraph 5.22 Log File Manager). During the data transfer, the display visualizes DUMP writing; to stop the data transfer, press ESC key on the instrument or on the PC.

7.1.2 Erase function: cancellation of the memorized data

To cancel the content of the memory, you must use Erase Log function (see the paragraph 5.2.5 Log file Manager). The instrument cancels the internal memory and, at the end of the operation, goes back to the normal visualization.

NOTES:

- The data transfer doesn't involve the memory cancellation: it's possible to repeat over and over the transfer process.
- The memorized data independently remain in memory from the batteries charge condition.
- To print the data using a printer provided with a parallel interface, it's necessary to use a serial- parallel converter (usually, not supplied).
- **The direct connection between the instrument and the printer with USB connection doesn't work.**
- During *logging*, some keys are disabled, **MEM**, **MENU**, **ENTER** and **ESC** are enabled.
- Pressing **MEM** and **MENU** keys, doesn't involve the memorized data if these keys are pressed **after** starting the memorization, vice versa it's valid what indicated below.

7.1.3 Print function

Pressing **ENTER** key, in real time, you can immediately send the detected data by the instrument to RS232C or USB ports. The measurement units of the printed data are the ones visualized on the display. The function starts, pressing **ENTER** key. The interval between the two following printings can be set between 15 seconds and 1 hour (see **Print interval** menu heading at the paragraph 5.3.2. Print Interval). If the printing interval is equal to 0, pressing **ENTER**, key you send only a data to the device. If the printing interval is more than 0, the data sending goes on till the operator doesn't stop it, pressing **ENTER** key again.


NOTE: During baud-rate setting, check the printing speed capacity of the used printer.

9. INSTRUMENT SIGNALS AND FAULTS

The following table lists all error indications and information displayed by the instrument and supplied to the user in different operating situations:

| Display indication | Explanation |
|-----------------------------------|--|
| ---.--- | This appears if the sensor relevant to the indicated physical quantity is not present or is faulty |
| OVFL | Overflow appears when the probe detects a value that exceeds the expected measurement range. |
| UFL | Underflow appears when the probe detects a lower value than the expected measurement range. |
| WARNING: MEMORY FULL!! | The instrument cannot store further data, the memory space is full. |
| LOG | Blinking message. It appears on the first line of the display and indicates a logging session. |

10. BATTERY SYMBOL AND BATTERY REPLACEMENT – MAIN POWER SUPPLY

The battery symbol 

on the display constantly shows the battery charge status. To the extent that batteries have discharged, the symbol "empties". When the charge decreases still further it starts blinking.



In this case, batteries should be replaced as soon as possible.


If you continue to use it, the instrument can no longer ensure correct measurement and turns off. Data stored on memory will remain.

The battery symbol becomes [≈] when the external power supply is connected.

To replace the batteries, go on as indicated below:

1. switch the instrument off;
2. disconnect the external power supply, if connected;
3. unscrew the battery cover counter clockwise and take out the battery holder.
4. replace the batteries (4 1.5V alkaline batteries – 1.5 V AA type). Check that the battery polarity matches the indication on the battery holder;
5. Replace the battery holder and screw the cover on clockwise.

The instrument can be powered by the main using, for example, the stabilized power supply SWD10 input 100÷240Vac output 12Vdc – 1000mA. The positive connector is in the middle.



The external diameter of power supply connector is 5.5mm, the internal diameter is 2.1mm.

Warning: The power supply cannot be used as battery charger. If the instrument is connected to the external power supply, the [≈] symbol is displayed instead the battery symbol.

Malfunctioning upon turning on after battery replacement

After replacing the batteries, the instrument may not restart correctly; in this case, repeat the operation.

After disconnecting the batteries, wait a few minutes in order to allow circuit condensers to discharge completely; then reinsert the batteries.

10.1 WARNING ABOUT THE BATTERIES USE

- Batteries should be removed when the instrument is not used for an extended time.
- Flat batteries must be replaced immediately.
- Avoid loss of liquid from batteries.
- Use waterproof and good-quality batteries, if possible alkaline. Sometimes on the market, it is possible to find new batteries with an insufficient charge capacity.

11. INSTRUMENT STORAGE

Instrument storage conditions:

- Temperature: -25...+65°C.
- Humidity: less than 90% RH without condensation.
- During storage avoid locations where:
 - humidity is high;
 - the instrument may be exposed to direct sunlight;
 - the instrument may be exposed to a source of high temperature;
 - the instrument may be exposed to strong vibrations;
 - the instrument may be exposed to steam, salt or any corrosive gas.

Some parts of the instrument are made of ABS plastic, polycarbonate: do not use any incompatible solvent for cleaning.

12. PRINTING OF THE MEASUREMENT REPORT

Evaluation Report

Hot Environments: Determination of WBGT heat stress index

Norm ISO 7243



Delta OHM
Via Marconi, 5
35030 Caselle di Selvazzano
Padova
Italy

INTRODUCTION

The WBGT (*Wet Bulb Globe Temperature*) (UNI, 1996) is an empirical temperature index used to evaluate very hot thermal environments, from the experimental correlations between microclimatic parameters and physiologic reactions of a large sample of subjects.

In order to determine the conditions of thermal stress within an environment, you should know the air temperature, wind speed, and air humidity, as well as the average radiation temperature. The WBGT index uses some derived quantities to characterise the environment under consideration from a thermal point of view.

PURPOSE AND APPLICATION SCOPE


The purpose of this survey is the WBGT index evaluation in a hot environment.

REFERENCE STANDARDS


Norm ISO 7243

NOTES


Space for notes

| | | |
|---|---|---------------------------------------|
|  | <p style="text-align: center;">Evaluation Report Hot Environments: Determination of WBGT heat stress index Norm ISO 7243</p> | <p>Mod. 001 rev.0 Page 3 of 6</p> |
| Instrumentation used: | | |
| Instrument Code: | Model HD32. | |
| Firmware Version: | Firm.Ver.=01.00 | |
| Firmware Date (yyyy/mm/dd): | Firm.Date=2005/10/12 | |
| Instrument Serial Number: | SN=12345678 | |
| User Code: | User ID=0000000000000000 | |
| Probes used: | | |
| Input description Ch.1 | | |
| Type of probe: | Pt100 | |
| Cal. Date: | 2004/09/13 | |
| Y/N: | 87654321 | |
| Input description Ch.2 | | |
| Type of probe: | Pt100 Tg 50 | |
| Cal. Date: | 2005/06/27 | |
| Y/N: | 05013380 | |
| Input description Ch.3 | | |
| Type of probe: | Pt100 Tw | |
| Cal. Date: | 2002/01/02 | |
| Y/N: | 04006422 | |

Warm Environments

| | | |
|--|--|--|
|  | <p style="text-align: center;">Evaluation Report Hot Environments: Determination of WBGT heat stress index Norm ISO 7243</p> | <p style="text-align: right;">Mod. 001 rev.0 Page 4 of 6</p> |
| Description of the observation location: | | |
| <p>Very Hot Environment Indoor, without solar radiation Person acclimatized to heat The worker being observed has an average size body</p> | | |
| Description of clothing: | | |
| <p>Daily Clothing: Intimate underwear and lingerie, short-sleeved vest/top, blouse, trousers, jacket, ankle socks, shoes</p> | | |
| <u>1.5 clo</u> | | |
| Description of activity: | | |
| <p>Type of Job: Sedentary activity (office, home, school, laboratory)</p> | | |
| <u>70 W/m2</u> | | |




Warm Environments

| | | |
|---|---|----------------|
|  | Evaluation Report | Mod. 001 rev.0 |
| | Hot Environments: Determination of WBGT heat stress index | Page 5 of 6 |
| | Norm ISO 7243 | |


Graph Trend:

Start date: 2006/10/05 **Start time:** 10:30:00
End date: 2006/10/05 **End time:** 10:38:00

Acquisition frequency: 15 sec



Warm Environments

| | | |
|---|--|--|
|  | <p style="text-align: center;">Evaluation Report Hot Environments: Determination of WBGT heat stress index Norm ISO 7243</p> | <p style="text-align: right;">Mod. 001 rev.0 Page 6 of 6</p> |
|---|--|--|

Measurements:

| | |
|--|--------|
| Globe Thermometer Temperature, T _g (°C) | 22 . 7 |
| Wet Bulb Temperature, T _w (°C) | 16 . 8 |
| Air Temperature, T _a (°C) | 22 . 7 |

Overall result:

| | |
|-----------------------------|--------|
| WBGT heat stress index (°C) | 18 . 6 |
| WBGT value limit (°C) | 28 . 0 |

13. NOTES ON WORKING AND OPERATING SAFETY

Authorized use

Comply with the technical specifications outlined in the chapter “TECHNICAL CHARACTERISTICS”. Its use is authorized only in conformity with the instructions written in this manual. Any different use is considered improper.

General instructions on security

This instrument has been manufactured and tested according to safety regulation EN 61010-1:2010 concerning electronic measurement instruments and was delivered ex factory in perfect security conditions.

Its regular functioning and operating security can be ensured only if all the normal safety measures as well as the specifications described in this manual are complied with.

Its regular functioning and operating security can be ensured only within the climatic conditions specified in the chapter “TECHNICAL CHARACTERISTICS”.

Do not use or store the instrument in ways and/or places in which there are:

- Quick environment temperature changes that could cause condensation.
- Corrosive or inflammable gases.
- Direct vibrations or shocks against the instrument.
- High intensity electromagnetic fields, static electricity.

If the instrument is moved from a cold to a hot environment, the condensation can disturb its functioning. In this case, you need to wait for the instrument to reach the environment temperature before using it.

User obligations

The user of the instrument must ensure that the following regulations and directives concerning the handling of hazardous materials are complied with:

- CEE directives on job safety
- National laws on job safety
- Accident prevention regulations

14. TECHNICAL CHARACTERISTICS

Instrument

| | |
|--------------------------------------|---|
| Dimensions (Length x Width x Height) | 185x90x40 mm |
| Weight | 470 g (complete of batteries) |
| Materials | ABS, rubber |
| Display | Dot matrix with backlight 160x160 points, visible area 52x42mm |

Working conditions

| | |
|---------------------------|------------------------------|
| Operative temperature | -5 ... 50°C |
| Storage temperature | -25 ... 65°C |
| Humidity relative to work | 0 ... 90% HR no condensation |

Protection degree

IP67

Instrument uncertainty

± 1 digit @ 20°C

Power supply

| | |
|---|---|
| Net power supply (code SWD10) | 12Vdc/1A |
| Batteries | 4 batteries 1.5V AA type |
| Autonomy | 200 hours with 1800mAh alkaline batteries |
| Absorbed current with switched instrument off | < 45µA |

Safety of the memorized data

unlimited

TP3207.2, TP3207 temperature probe

| | |
|---------------------------------|---|
| Sensor type: | Thin film Pt100 |
| Accuracy: | Class 1/3 DIN |
| Measurement range: | -40 ÷ 100 °C |
| Resolution: | 0.1°C |
| Drifting in temperature @20°C: | 0.003%/°C |
| Drifting after 1 year: | 0.1°C/year |
| Connection: | 4 wires plus SICRAM module |
| Connector: | 8 – poles female DIN45326 |
| Cable: | Only TP3207 (2m) |
| Dimensions: | Ø=14 mm L= 150 mm (TP3207.2), L= 140 mm (TP3207) |
| Response time T ₉₅ : | 15 minutes |

Globe thermometer probe Ø=50 mm TP3276.2, Ø=150 mm TP3275

| | |
|---------------------------------|--|
| Sensor type: | Pt100 |
| Accuracy: | Class 1/3 DIN |
| Measurement range: | -10 ÷ 100 °C |
| Resolution: | 0.1°C |
| Drifting in temperature @20°C: | 0.003%/°C |
| Drifting after 1 year: | 0.1°C/year |
| Connection: | 4 wires plus SICRAM module |
| Connector: | 8 – poles female DIN45326 |
| Cable: | Only TP3275 (2m) |
| Stem dimensions: | Ø=8 mm L= 170 mm (TP3276.2), Ø=14 mm L= 110 mm (TP3275) |
| Response time T ₉₅ : | 15 minutes |

HP3201.2, HP3201 Natural ventilation wet bulb probe

| | |
|---------------------------------|---|
| Sensor type: | Pt100 |
| Accuracy: | Class A |
| Measurement range: | 4 °C ÷ 80 °C |
| Resolution: | 0.1°C |
| Drifting in temperature @20°C: | 0.003%/°C |
| Drifting after 1 year: | 0.1°C/year |
| Connection: | 4 wires plus SICRAM module |
| Connector: | 8 – poles female DIN45326 |
| Cable: | Only HP3201 (2m) |
| Stem dimensions: | Ø=14 mm L= 170 mm (HP3201.2), L= 110 mm (HP3201) |
| Cotton wick length: | about 10 cm. |
| Tank capacity: | 15 cc. |
| Tank autonomy: | 96 hours with RH=50%, t = 23°C |
| Response time T ₉₅ : | 15 minutes |

TP3204S Natural ventilation wet bulb probe

| | |
|---------------------------------|---|
| Sensor type: | Pt100 |
| Accuracy: | Class A |
| Measurement range: | 4 °C ÷ 80 °C |
| Resolution: | 0.1°C |
| Drifting in temperature @20°C: | 0.003%/°C |
| Drifting after 1 year: | 0.1°C/year |
| Connection: | 4 wires plus SICRAM module |
| Connector: | 8 – poles female DIN45326 |
| Cable: | 2 m |
| Dimensions: | L x W x H=140 x 65 x 178.5 mm (reservoir + bottle) |
| Scotch length: | about 10 cm |
| Tank capacity: | 500 cc |
| Tank autonomy: | 15 days @ t = 40 °C |
| Response time T ₉₅ : | 15 minutes |

Connections

Inputs for probes with SICRAM module 3 male 8-pole DIN 45326 connectors

Serial interface:

| | |
|---------------|--|
| Pin: | M12-8 poles. |
| Type: | RS232C (EIA/TIA574) or USB 1.1 o 2.0 not isolated |
| Baud rate: | from 1200 to 38400 baud. with USB baud=460800 |
| Data Bit: | 8 |
| Parity: | None |
| Stop bit: | 1 |
| Flow control: | Xon-Xoff |
| Cable length: | max 15m |

Memory

divided in 64 blocks.

Memory capacity

67600 memorizations for each of 3 inputs

Memorization interval

selectable between: 15, 30 seconds, 1, 2,
5, 10, 15, 20, 30 minutes and 1 hour

15. ORDERING CODES

HD32.2 Kit consists of:

- **HD32.2 WBGT Index instrument**, 4 alkaline batteries 1.5V AA type, instruction manual, case. **The probes and the cables are not included.**
- **DeltaLog10 Software for warm environments: WBGT analysis.**
The DeltaLog10 software is downloadable from Delta OHM website.

Necessary probes for **WBGT** measurement:

- **TP3207.2** Dry bulb temperature probe.
- **TP3276.2** Globe thermometer probe.
- **HP3201.2** Wet bulb temperature probe with natural ventilation.

Necessary probes for **WBGT version A** measurement:

- **TP3207** Dry bulb temperature probe.
- **TP3275** Globe thermometer probe.
- **HP3201** or **TP3204S** Wet bulb temperature probe with natural ventilation.

15.1 PROBES FOR HD32.2 WBGT INDEX

| | |
|-----------------|---|
| TP3207.2 | Temperature probe for Pt100 sensor. Shank probe Ø 14mm, length 150mm. Complete with SICRAM module. |
| TP3276.2 | Globe thermometer probe for Pt100 sensor, globe Ø 50 mm. Shank Ø 8 mm, length 170 mm. Complete with SICRAM module. |
| HP3201.2 | Wet bulb probe with natural ventilation. Pt100 sensor. Shank probe Ø 14 mm, length 170 mm complete with SICRAM module, spare parts of the cotton wick and case of 50cc. distilled water. |
| TP3207 | Temperature probe for Pt100 sensor. Shank probe Ø 14mm, length 140mm. Cable length 2 m. Complete with SICRAM module. |
| TP3275 | Globe thermometer probe for Pt100 sensor, globe Ø 150 mm. Shank Ø 14 mm, length 110 mm. Cable length 2 m. Complete with SICRAM module. |
| HP3201 | Natural ventilation wet bulb probe. Pt100 sensor. Probe stem Ø 14 mm, length 110 mm. Cable length 2 m. Complete with SICRAM module, spare parts of the cotton wick and and case of 50 cc of distilled water. |
| TP3204S | Natural ventilation wet bulb probe for long-lasting measurements. Capacity: 500 cc of distilled water. Pt100 sensor. Cable length 2 m. Complete with SICRAM module, 500 cc bottle and two spare cotton wicks. |

Accessories:

| | |
|----------------------|---|
| VTRAP30 | Tripod to be fixed to the instrument with a maximum height of 280 mm |
| VTRAP32.2A.3A | Tripod for HD32.2A |
| HD32.2.7.1 | Support for probes, to be fixed on standard tripod for version HD32.2A |
| HD2110/RS | Connection cable with M12 connector on the instrument side and with |

| | |
|-------------------|--|
| | 9-pole SubD female connector for RS232C on PC side. |
| HD2110/USB | Connection cable with M12 connector on the instrument, USB 2.0 connector on PC side. |
| SWD10 | Stabilized power supply with 100-240Vac/12Vdc-1A main tension |
| AQC | 200cc of distilled water |
| HD40.1 | Printer (it uses HD2110/RS cable) |

DELTA OHM metrology laboratories LAT N° 124 are ISO/IEC 17025 accredited by ACCREDIA for Temperature, Humidity, Pressure, Photometry / Radiometry, Acoustics and Air Velocity. They can supply calibration certificates for the accredited quantities.

NOTES

**DICHIARAZIONE DI CONFORMITÀ UE
EU DECLARATION OF CONFORMITY****Delta Ohm S.r.L. a socio unico – Via Marconi 5 – 35030 Caselle di Selvazzano – Padova – ITALY**Documento Nr. / Mese.Anno: **5087 / 11.2017**
Document-No. / Month.Year :

Si dichiara con la presente, in qualità di produttore e sotto la propria responsabilità esclusiva, che i seguenti prodotti sono conformi ai requisiti di protezione definiti nelle direttive del Consiglio Europeo:

We declare as manufacturer herewith under our sole responsibility that the following products are in compliance with the protection requirements defined in the European Council directives:

Codice prodotto: **HD32.2**
Product identifier :Descrizione prodotto: **Analizzatore dell'indice microclimatico WBGT**
Product description : **WBGT microclimatic index analyzer**I prodotti sono conformi alle seguenti Direttive Europee:
The products conform to following European Directives:

| Direttive / Directives | |
|------------------------|--|
| 2014/30/EU | Direttiva EMC / EMC Directive |
| 2014/35/EU | Direttiva bassa tensione / Low Voltage Directive |
| 2011/65/EU | RoHS / RoHS |

Norme armonizzate applicate o riferimento a specifiche tecniche:
Applied harmonized standards or mentioned technical specifications:

| Norme armonizzate / Harmonized standards | |
|--|---|
| EN 61010-1:2010 | Requisiti di sicurezza elettrica / Electrical safety requirements |
| EN 61326-1:2013 | Requisiti EMC / EMC requirements |
| EN 50581:2012 | RoHS / RoHS |

Il produttore è responsabile per la dichiarazione rilasciata da:
The manufacturer is responsible for the declaration released by:

Johannes Overhues

Amministratore delegato
Chief Executive Officer

Caselle di Selvazzano, 24/11/2017

Questa dichiarazione certifica l'accordo con la legislazione armonizzata menzionata, non costituisce tuttavia garanzia delle caratteristiche.

This declaration certifies the agreement with the harmonization legislation mentioned, contained however no warranty of characteristics.

GUARANTEE



Member of GHM GROUP

TERMS OF GUARANTEE

All DELTA OHM instruments are subject to accurate testing, and are guaranteed for 24 months from the date of purchase. DELTA OHM will repair or replace free of charge the parts that, within the warranty period, shall be deemed non efficient according to its own judgement. Complete replacement is excluded and no damage claims are accepted. The DELTA OHM guarantee only covers instrument repair. The guarantee is void in case of incidental breakage during transport, negligence, misuse, connection to a different voltage than that required for the appliance by the operator. Finally, a product repaired or tampered by unauthorized third parties is excluded from the guarantee. The instrument shall be returned FREE OF SHIPMENT CHARGES to your dealer. The jurisdiction of Padua applies in any dispute.



The electrical and electronic equipment marked with this symbol cannot be disposed of in public landfills. According to the Directive 2011/65/EU, the european users of electrical and electronic equipment can return it to the dealer or manufacturer upon purchase of a new one. The illegal disposal of electrical and electronic equipment is punished with an administrative fine.

This guarantee must be sent together with the instrument to our service centre.

IMPORTANT: Guarantee is valid only if coupon has been correctly filled in all details.

Instrument Code: **HD32.2**

Serial Number _____

RENEWALS

Date _____

Date _____

Inspector _____

Inspector _____

Date _____

Date _____

Inspector _____

Inspector _____

Date _____

Date _____

Inspector _____

Inspector _____



The quality level of our instruments is the result of the constant development of the product. This may produce some differences between the information written in this manual and the instrument you have purchased. We cannot completely exclude the possibility of errors in the manual, for which we apologize.

The data, images and descriptions included in this manual cannot be legally asserted. We reserve the right to make changes and corrections with no prior notice.

V1.4
25/07/2018