



EU DECLARATION OF CONFORMITY

Product Identification:

First declared: 2006

Fluke Networks 3000Pro, 3000Pro F50, 3000Pro F60 - Analog Toner and Probe

Statement of Conformity:

This Declaration of Conformity is issued under the sole responsibility of the manufacturer, based on sample product type-test, using ISO 9001:2015 Quality Management System. The manufacturer hereby declares this product conforms to the following EU Directives:

Directive 2011/65/EU, RoHS

Directive 2014/30/EU, Electromagnetic Compatibility (EMC)

Standards Used:

EN 50581: 2012 Restriction of the use of hazardous substances

EN 61326-1: 2013 Electrical equipment for measurement, control and laboratory-EMC; General

EN 61010-1: 2010 Safety requirements for measurement, control and laboratory; General

Special Conditions:

None

Manufacturer

Fluke Networks
6920 Seaway Boulevard M/S 143F
Everett, WA 98203; USA

DC20180425131010

Thomas M. Smith P.E.

Product Compliance Manager, Fluke Corporation

DC20180425131010 - This declaration supersedes all previous declarations for this product.

25 April 2018