



FLIR COMMERCIAL SYSTEMS, INC. - EXTECH BRAND
9 Townsend West, Nashua NH 03063 / Phone: 603.324.7800 / Fax: 603.324.7864

Declaration of Conformity

Extech Model: IR200

Description: Non-Contact Forehead InfraRed Thermometer

Date of Issue: 08-Oct-14

We, FLIR Commercial Systems, Inc. - Extech Brand, 9 Townsend West, Nashua, NH 03063 declare that a sample of the product listed above has been tested by a third party for CE marking according to:

Medical Directive: 93/42/EEC
Report Number: 17040346 001
Report Date: 9/15/2014

Standards:


IEC 60601-1:1988+A1:1991+A2:1995
EN 60601-1:1990+A1:1993+A2:1995
EN ISO 13485:2012
EN ISO 13485:2012/AC:2012

RoHS Directive: 2011/65/EU

Standard:

EN 50581:2012

The test reports show that the product fulfills the requirement in the EC Medical Directive and RoHS Directive for CE Marking. On this basis, together with the manufacturer's own documented production control, the manufacturer (or his European authorized representative) can in his EC Declaration of Conformity verify compliance with the EC Medical Directive and RoHS Directive.



Tony Campagna
Director of Quality Services