

# Kalibrier-Zertifikat Calibration certificate

MUSTER

Gegenstand Object	Temperatur-Messgerät testo 925	Tauch-/Einstechfühler was- serdicht
Hersteller Manufacturer	Testo AG	Testo AG
Typ Type description	0560 9250	0602 1293
Serien Nr. Serial no.	12345678	12345678
Inventar Nr. Inventory no.	---	---
Prüfmittel Nr. Test equipment no.	---	---
Equipment Nr. Equipment no.	12345678	12345678
Standort Location	---	---
Auftraggeber Customer	Max Mustermann GmbH DE-12345 Musterstadt	
Kunden Nr. Customer ID no.	1234567	
Auftrags Nr. Order no.	1234567 / 0520 0001	
Datum der Kalibrierung Date of calibration	04.04.2014	
Datum der empfohlenen Rekalibrierung Date of the recommended re-calibration	04.04.2015	

## Konformitätsaussage Conformity

- Messwert(e) innerhalb der zulässigen Abweichung<sup>1</sup>. Measured value(s) within the allowable deviation<sup>1</sup>.  
 Messwert(e) außerhalb der zulässigen Abweichung<sup>1</sup>. Measured value(s) outside of the allowable deviation<sup>1</sup>.

--- ---

--- ---

<sup>1)</sup> Die Messunsicherheit wurde nach GUM mit dem Erweiterungsfaktor k=2 berechnet und enthält die Unsicherheit des Verfahrens sowie die Unsicherheit des Prüflings. Die Konformitätsaussage erfolgte nach DIN EN ISO 14253-1 gemäß der Kalibrieranweisung QSA-TIS 7.5-02.

<sup>1)</sup> The measurement uncertainty was calculated according to the regulations of GUM with the coverage factor k=2 and contains the uncertainty of the measuring procedure and the uncertainty of the measuring system. The statement of conformity was made according to DIN EN ISO 14253-1 according to calibration instruction QSA-TIS 7.5-02.

Dieser Kalibrierschein darf nur vollständig weiterverbreitet werden. Auszüge oder Änderungen bedürfen der Genehmigung des ausstellenden Kalibrierlaboratoriums. Kalibrierscheine ohne Unterschrift und Stempel haben keine Gültigkeit.

This calibration certificate may not be reproduced other than in full except with the permission of the issuing laboratory. Calibration certificates without signature and seal are not valid.

Stempel Seal



Fachverantwortlicher Supervisor

  
Max Mustermann

Bearbeiter Technician

  
Max Mustermann

# Kalibrier-Zertifikat Calibration certificate

MUSTER

## Messeinrichtungen Measuring equipment

Index	Referenz Reference	Rückführung Traceability	Rekal. Next cal.	Zertifikat-Nr. Certificate-no.	Eq.-Nr. Eq.-no.
a	testo 781 mit Fühler testo 781 with probe	15070-01-01 2014-03	2015-03	T52209	10211690
b	testo 781 mit Fühler testo 781 with probe	15070-01-01 2014-01	2015-01	T50822	10211689
c	testo 781 mit Fühler testo 781 with probe	15070-01-01 2014-01	2015-01	T50821	10211686

Referenzzertifikate sind auf [www.primasonline.de](http://www.primasonline.de) abrufbar Reference certificates are available at [www.primasonline.com](http://www.primasonline.com)

## Umgebungsbedingungen Ambient conditions

Temperatur Temperature	(20...26) °C	Druck Pressure	(940...990) hPa
Feuchte Humidity	(20...60) % rF % RH	Luftdichte Air density	--- kg/m³

## Messverfahren Measuring procedure

Vergleichsmessung in umgewälzten Flüssigkeitsbädern.  
Measuring in a circulated liquid bath.

## Messergebnisse Measuring results

Kanal Channel ---

Messwert Referenz Measured value reference	Angezeigter Messwert Kalibriergegenstand Indicated measured value probe	Abweichung Deviation	Zulässige Abweichung <sup>2)</sup> Allowed deviation <sup>2)</sup>	Messunsicherheit (k=2) Measuring uncertainty (k=2)	Bewertung Confirmation
°C	°C	°C	°C	°C	
-18,16 <sup>a</sup>	-17,4	0,76	± 3,2	0,14	pass
-0,05 <sup>b</sup>	0,1	0,15	± 3,1	0,14	pass
60,04 <sup>c</sup>	59,1	-0,94	± 3,3	0,14	pass

<sup>2)</sup> gemäß Hersteller in accordance with the manufacturer

## Besondere Bemerkungen Special remarks

---

---

---

---

---

---